

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
 Data File : VU047140.D
 Acq On : 14 Feb 2022 18:54
 Operator : SY/MD
 Sample : N1205-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-UNRVOA-WS

Quant Time: Feb 15 03:53:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	54166	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	20227	0.964	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	21397	0.979	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.000%	
Target Compounds						
					Qvalue	
5) Bromomethane	1.864	94	108380	9.060	ug/l	99
6) Chloroethane	1.941	64	175493	16.682	ug/l	100
10) Iodomethane	2.732	142	16006	1.856	ug/l	99
17) 1,1-Dichloroethane	3.877	63	391832	15.827	ug/l	98
21) 2,2-Dichloropropane	4.671	77	236988	10.711	ug/l	98
25) Methyl tert-Butyl Ether	3.372	73	708037	19.720	ug/l	100
26) Bromochloromethane	4.983	128	25578	3.660	ug/l	88
29) 1,1-Dichloropropene	5.529	75	359537	18.059	ug/l	99
43) Dibromomethane	6.922	93	148120	18.603	ug/l	91
50) t-1,3-Dichloropropene	8.214	75	295008	14.473	ug/l	99
51) cis-1,3-Dichloropropene	7.610	75	420259	18.392	ug/l	98
53) 1,3-Dichloropropane	8.578	76	359250	17.700	ug/l	100
60) 1,1,1,2-Tetrachloroethane	9.536	131	260113	18.861	ug/l	99
69) Isopropylbenzene	10.487	105	914177	13.051	ug/l	98
70) 1,1,2,2-Tetrachloroethane	10.783	83	280694	18.276	ug/l	99
71) 1,2,3-Trichloropropane	10.825	75	208941	16.119	ug/l	85
72) Bromobenzene	10.783	156	63059	3.686	ug/l	88
73) n-propylbenzene	10.909	120	365330	18.945	ug/l	87
74) 2-Chlorotoluene	10.989	126	141876	8.630	ug/l	91
75) 1,3,5-Trimethylbenzene	11.089	105	448112	7.801	ug/l	99
76) 4-Chlorotoluene	11.098	126	301187	17.438	ug/l	75
77) tert-Butylbenzene	11.423	119	286504	4.970	ug/l	97
78) 1,2,4-Trimethylbenzene	11.468	105	879070	14.995	ug/l	97
79) sec-Butylbenzene	11.645	105	864332	11.292	ug/l	98
81) p-Isopropyltoluene	11.796	119	1240642	19.155	ug/l	97
82) 1,3-Dichlorobenzene	11.748	146	572507	17.190	ug/l	98
84) n-Butylbenzene	12.211	91	515795	8.293	ug/l	97
88) Hexachlorobutadiene	14.024	225	427324	33.206	ug/l	98
89) Naphthalene	14.085	128	1917802	37.197	ug/l	99
90) 1,2,3-Trichlorobenzene	14.330	180	1088438	48.209	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
Data File : VU047140.D
Acq On : 14 Feb 2022 18:54
Operator : SY/MD
Sample : N1205-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-UNRVOA-WS

Quant Time: Feb 15 03:53:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

