

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
 Data File : VU047132.D
 Acq On : 14 Feb 2022 14:39
 Operator : SY/MD
 Sample : VU0214WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0214WBL01

Quant Time: Feb 15 03:48:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	56250	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	19965	0.916	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	20399	0.899	ug/l	0.00
Spiked Amount 1.000			Recovery	=	90.000%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021422\
Data File : VU047132.D
Acq On : 14 Feb 2022 14:39
Operator : SY/MD
Sample : VU0214WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0214WBL01

Quant Time: Feb 15 03:48:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.118 min Scan# 1486
Delta R.T. -0.000 min
Lab File: VU047132.D
Acq: 14 Feb 2022 14:39




Instrument :
MSVOA_U
ClientSampleId :
VU0214WBL01


Tgt Ion: 96	Resp: 56250
Ion Ratio	Lower Upper
96 100	
70 19.4	16.8 25.2
95 8.9	7.3 10.9
97 6.5	0.0 0.0#



#57
4-Bromofluorobenzene
Concen: 0.916 ug/l
RT: 10.635 min Scan# 2891
Delta R.T. -0.000 min
Lab File: VU047132.D
Acq: 14 Feb 2022 14:39

Tgt Ion: 95	Resp: 19965
Ion Ratio	Lower Upper
95 100	
174 83.1	56.6 84.8
176 80.8	54.3 81.5





#68
 1,2-Dichlorobenzene-d4
 Concen: 0.899 ug/l
 RT: 12.195 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: VU047132.D
 Acq: 14 Feb 2022 14:39

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0214WBL01

Tgt Ion	Ratio	Lower	Upper
152	100		
115	65.5	0.0	251.0
150	160.5	0.0	613.8

