

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021522\
 Data File : VU047147.D
 Acq On : 15 Feb 2022 12:30
 Operator : SY/MD
 Sample : N1205-10DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WSDL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/18/2022
 Supervised By : Mahesh Dadoda 02/18/2022

Quant Time: Feb 17 08:03:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	52337	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	19459	0.960	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	20550	0.973	ug/l	0.00
Spiked Amount 1.000			Recovery	=	97.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	71468	5.946	ug/l	94
24) tert-Butyl Alcohol	3.234	59	5177m	7.962	ug/l	
25) Methyl tert-Butyl Ether	3.372	73	322692	9.302	ug/l	99
31) Isopropyl Ether	4.002	45	59214	1.548	ug/l	97
32) Ethyl-t-butyl ether	4.510	59	261065	6.712	ug/l	98
33) Tert-Amyl methyl ether	5.948	73	356054	9.939	ug/l	98
73) n-propylbenzene	10.909	120	178426	9.576	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021522\
Data File : VU047147.D
Acq On : 15 Feb 2022 12:30
Operator : SY/MD
Sample : N1205-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/18/2022
Supervised By : Mahesh Dadoda 02/18/2022

Quant Time: Feb 17 08:03:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

