

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021522\
 Data File : VU047150.D
 Acq On : 15 Feb 2022 15:12
 Operator : SY/MD
 Sample : N1205-07DL 2X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-RVOA-WSDL

Quant Time: Feb 17 08:04:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	53631	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	20294	0.977	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	21196	0.980	ug/l	0.00
Spiked Amount 1.000			Recovery	=	98.000%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.610	62	130490	7.197	ug/l	99
9) 1,1-Dichloroethene	2.591	96	71735	5.659	ug/l	95
15) Methylene Chloride	3.054	84	51556	3.621	ug/l	97
16) trans-1,2-Dichloroethene	3.359	96	52718	3.940	ug/l	96
22) cis-1,2-Dichloroethene	4.674	96	55028	3.576	ug/l	84
28) 1,1,1-Trichloroethane	5.324	97	128934	5.988	ug/l	97
30) Carbon Tetrachloride	5.529	117	57348	3.029	ug/l	98
35) Benzene	5.780	78	315170	5.705	ug/l	99
36) 1,2-Dichloroethane	5.800	62	169605	10.141	ug/l	99
37) Trichloroethene	6.545	130	74559	4.875	ug/l	96
38) 1,2-Dichloropropane	6.796	63	136432	9.333	ug/l	100
49) Toluene	7.973	92	294746	8.272	ug/l	98
52) 1,1,2-Trichloroethane	8.404	97	81761	7.269	ug/l	98
58) Tetrachloroethene	8.555	164	66701	4.726	ug/l	98
59) Chlorobenzene	9.449	112	198680	5.001	ug/l	100
63) Ethyl Benzene	9.574	91	250043	3.580	ug/l	98
64) m/p-Xylenes	9.697	106	152821	5.740	ug/l	98
65) o-Xylene	10.102	106	150859	5.850	ug/l	95
66) Styrene	10.118	104	393404	8.988	ug/l	94
83) 1,4-Dichlorobenzene	11.838	146	107402	3.233	ug/l	98
85) 1,2-Dichlorobenzene	12.214	146	94494	2.977	ug/l	97
87) 1,2,4-Trichlorobenzene	13.844	180	53233	2.136	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021522\
Data File : VU047150.D
Acq On : 15 Feb 2022 15:12
Operator : SY/MD
Sample : N1205-07DL 2X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-RVOA-WSDL

Quant Time: Feb 17 08:04:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

