

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
 Data File : VU053105.D
 Acq On : 15 Feb 2023 10:54
 Operator : JC/MD
 Sample : VSTD0.0529
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05029

Quant Time: Feb 16 00:44:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 16 00:43:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	38665	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.349	117	33930	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	19220	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	1321	0.051	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.661	65	1545	0.064	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	1323	0.046	ug/L	0.00
8) Toluene-d8	7.833	98	4160	0.058	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.682	84	1270	0.051	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.598	62	1656	0.061	ug/L #	69
6) Trichloroethene	6.489	95	1846	0.061	ug/L	91
9) 1,2-Dibromoethane	8.854	107	838	0.045	ug/L #	95
12) 1,2,3-Trichloropropane	10.751	75	685	0.042	ug/L	95
13) 1,2-Dibromo-3-chloropr...	12.927	75	158	0.046	ug/L #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
Data File : VU053105.D
Acq On : 15 Feb 2023 10:54
Operator : JC/MD
Sample : VSTD0.0529
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05029

Quant Time: Feb 16 00:44:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 16 00:43:36 2023
Response via : Initial Calibration

