

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
 Data File : VU053110.D
 Acq On : 15 Feb 2023 13:32
 Operator : JC/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV034

Quant Time: Feb 16 00:58:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 16 00:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	37871	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.346	117	33016	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	18850	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	12681	0.500	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.000%
4) 1,2-Dichloroethane-d4	5.661	65	10972	0.462	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	6.639	67	14779	0.526	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.830	98	33672	0.479	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	12291	0.511	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.598	62	12431	0.464	ug/L	100
6) Trichloroethene	6.489	95	14070	0.476	ug/L	100
9) 1,2-Dibromoethane	8.866	107	9553	0.522	ug/L	100
12) 1,2,3-Trichloropropane	10.748	75	8292	0.522	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.924	75	1795	0.529	ug/L #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
Data File : VU053110.D
Acq On : 15 Feb 2023 13:32
Operator : JC/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV034

Quant Time: Feb 16 00:58:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 16 00:50:34 2023
Response via : Initial Calibration

