

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
 Data File : VU053106.D
 Acq On : 15 Feb 2023 11:42
 Operator : JC/MD
 Sample : VSTD0.130
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1030

Quant Time: Feb 16 00:45:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 16 00:43:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	37507	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.345	117	32138	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.741	152	17862	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	2816	0.112	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.660	65	2531	0.108	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	2597	0.095	ug/L	0.00
8) Toluene-d8	7.830	98	7254	0.106	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.679	84	2217	0.095	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.598	62	2920	0.110	ug/L	95
6) Trichloroethene	6.489	95	2790	0.097	ug/L	95
9) 1,2-Dibromoethane	8.866	107	1754	0.098	ug/L #	97
12) 1,2,3-Trichloropropane	10.747	75	1505	0.100	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.924	75	340	0.106	ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021523\
Data File : VU053106.D
Acq On : 15 Feb 2023 11:42
Operator : JC/MD
Sample : VSTD0.130
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1030

Quant Time: Feb 16 00:45:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM021523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 16 00:43:36 2023
Response via : Initial Calibration

