

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
 Data File : VU047164.D
 Acq On : 16 Feb 2022 16:27
 Operator : SY/MD
 Sample : N1558-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M74

Quant Time: Feb 17 04:06:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 13:52:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	180871	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	182947	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	105392	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	75192	4.922	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.400%
7) Chloroethane-d5	1.919	69	60363	4.986	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.800%
11) 1,1-Dichloroethene-d2	2.575	65	29316	4.640	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.800%
20) 2-Butanone-d5	4.643	46	209226	57.303	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.600%
24) Chloroform-d	5.067	84	125951	4.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.707	65	69551	4.929	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.600%
32) Benzene-d6	5.732	84	265483	5.194	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.800%
36) 1,2-Dichloropropane-d6	6.694	67	84689	5.207	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.900	98	255603	5.112	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
43) trans-1,3-Dichloroprop...	8.183	79	35385	4.799	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.000%
46) 2-Hexanone-d5	8.636	63	189179	48.728	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.460%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	70957	4.800	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	99484	5.261	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%
Target Compounds						Qvalue
21) 2-Butanone	4.726	43	15607	4.547	ug/L	97
42) Toluene	7.974	91	37120	0.641	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
Data File : VU047164.D
Acq On : 16 Feb 2022 16:27
Operator : SY/MD
Sample : N1558-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M74

Quant Time: Feb 17 04:06:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 16 13:52:04 2022
Response via : Initial Calibration

