

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
 Data File : VU047170.D
 Acq On : 16 Feb 2022 18:51
 Operator : SY/MD
 Sample : N1558-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M61

Quant Time: Feb 17 04:07:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 13:52:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	179227	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	181568	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	101739	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	69493	4.591	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.800%
7) Chloroethane-d5	1.919	69	56059	4.673	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.575	65	27745	4.432	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.600%
20) 2-Butanone-d5	4.642	46	193303	53.428	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.860%
24) Chloroform-d	5.070	84	118232	4.702	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	5.707	65	66078	4.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
32) Benzene-d6	5.732	84	248908	4.906	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	6.694	67	79281	4.911	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%
41) Toluene-d8	7.899	98	236020	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	8.182	79	32744	4.475	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.636	63	172132	44.674	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.340%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	66286	4.518	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	91650	5.021	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.400%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.607	62	25311	1.524	ug/L	99
22) cis-1,2-Dichloroethene	4.675	96	2627	0.189	ug/L	87
30) Cyclohexane	5.395	56	28866	1.345	ug/L	98
33) Benzene	5.781	78	9681	0.188	ug/L	100
52) Ethylbenzene	9.575	91	10795	0.172	ug/L	99
53) m,p-Xylene	9.694	106	17447	0.706	ug/L	97
54) o-Xylene	10.102	106	12776	0.531	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
Data File : VU047170.D
Acq On : 16 Feb 2022 18:51
Operator : SY/MD
Sample : N1558-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M61

Quant Time: Feb 17 04:07:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 16 13:52:04 2022
Response via : Initial Calibration

