

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
 Data File : VU047159.D
 Acq On : 16 Feb 2022 14:21
 Operator : SY/MD
 Sample : VU0216WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK065

Quant Time: Feb 16 14:41:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 13:52:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	184070	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	187991	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	109610	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	79116	5.089	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.800%
7) Chloroethane-d5	1.919	69	62731	5.092	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.800%
11) 1,1-Dichloroethene-d2	2.572	65	32657	5.079	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.600%
20) 2-Butanone-d5	4.645	46	206757	55.643	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.280%
24) Chloroform-d	5.070	84	131685	5.099	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.706	65	73207	5.098	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
32) Benzene-d6	5.732	84	281692	5.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.200%
36) 1,2-Dichloropropane-d6	6.694	67	88633	5.303	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.903	98	267647	5.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.200%
43) trans-1,3-Dichloroprop...	8.182	79	37654	4.970	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.400%
46) 2-Hexanone-d5	8.636	63	192532	48.261	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.520%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	72074	4.744	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	102822	5.228	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
Data File : VU047159.D
Acq On : 16 Feb 2022 14:21
Operator : SY/MD
Sample : VU0216WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK065

Quant Time: Feb 16 14:41:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 16 13:52:04 2022
Response via : Initial Calibration

