

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
 Data File : VU047173.D
 Acq On : 16 Feb 2022 20:03
 Operator : SY/MD
 Sample : N1558-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M64

Quant Time: Feb 17 04:08:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 13:52:04 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	177228	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	179237	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	101922	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	65025	4.344	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.800%	
7) Chloroethane-d5	1.919	69	53860	4.540	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.800%	
11) 1,1-Dichloroethene-d2	2.575	65	26025	4.204	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.000%	
20) 2-Butanone-d5	4.642	46	190028	53.115	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.240%	
24) Chloroform-d	5.067	84	116637	4.691	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.800%	
26) 1,2-Dichloroethane-d4	5.706	65	64409	4.659	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
32) Benzene-d6	5.732	84	241061	4.814	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
36) 1,2-Dichloropropane-d6	6.693	67	75991	4.768	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.400%	
41) Toluene-d8	7.902	98	227611	4.646	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
43) trans-1,3-Dichloroprop...	8.182	79	31988	4.428	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.600%	
46) 2-Hexanone-d5	8.636	63	168604	44.328	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.660%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	64695	4.467	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	88439	4.836	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.800%	
Target Compounds						
17) Methyl tert-butyl Ether	3.369	73	12734	0.400	ug/L	94
22) cis-1,2-Dichloroethene	4.671	96	16268	1.186	ug/L	98
34) Trichloroethene	6.546	95	12216	0.899	ug/L	95
47) Tetrachloroethene	8.555	164	73646	6.680	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
Data File : VU047173.D
Acq On : 16 Feb 2022 20:03
Operator : SY/MD
Sample : N1558-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M64

Quant Time: Feb 17 04:08:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 16 13:52:04 2022
Response via : Initial Calibration









#17
Methyl tert-butyl Ether
Concen: 0.400 ug/L
RT: 3.369 min Scan# 61
Delta R.T. -0.000 min
Lab File: VU047173.D
Acq: 16 Feb 2022 20:03




Instrument :
MSVOA_U
ClientSampleId :
F4M64


Tgt Ion: 73 Resp: 12734
Ion Ratio Lower Upper
73 100
43 23.6 14.8 22.2#
57 22.0 17.4 26.0



#22
cis-1,2-Dichloroethene
Concen: 1.186 ug/L
RT: 4.671 min Scan# 1036
Delta R.T. -0.003 min
Lab File: VU047173.D
Acq: 16 Feb 2022 20:03

Tgt Ion: 96 Resp: 16268
Ion Ratio Lower Upper
96 100
61 127.3 88.5 164.3
98 68.0 44.4 82.5









#34
Trichloroethene
Concen: 0.899 ug/L
RT: 6.546 min Scan# 1
Delta R.T. -0.000 min
Lab File: VU047173.D
Acq: 16 Feb 2022 20:03

Instrument :
MSVOA_U
ClientSampleId :
F4M64

Tgt Ion: 95 Resp: 12216

Ion	Ratio	Lower	Upper
95	100		
97	64.7	46.5	86.5
132	98.1	65.1	120.9
130	103.7	67.3	124.9



#47
Tetrachloroethene
Concen: 6.680 ug/L
RT: 8.555 min Scan# 2244
Delta R.T. -0.000 min
Lab File: VU047173.D
Acq: 16 Feb 2022 20:03

Tgt Ion: 164 Resp: 73646

Ion	Ratio	Lower	Upper
164	100		
129	92.2	71.8	133.4
131	90.4	68.5	127.3
166	125.9	94.1	174.8

