

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
 Data File : VU047178.D
 Acq On : 17 Feb 2022 00:03
 Operator : SY/MD
 Sample : N1559-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M51

Quant Time: Feb 17 04:52:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 16 13:52:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	171884	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	173969	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	99266	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	65014	4.479	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.600%
7) Chloroethane-d5	1.919	69	53059	4.612	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.575	65	25807	4.298	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.000%
20) 2-Butanone-d5	4.645	46	186539	53.761	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.520%
24) Chloroform-d	5.067	84	112438	4.663	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.706	65	61677	4.600	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
32) Benzene-d6	5.732	84	234526	4.825	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
36) 1,2-Dichloropropane-d6	6.694	67	75019	4.850	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.000%
41) Toluene-d8	7.902	98	223493	4.700	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	8.182	79	30118	4.296	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.636	63	157606	42.691	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.380%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	61267	4.358	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	86789	4.873	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021622\
Data File : VU047178.D
Acq On : 17 Feb 2022 00:03
Operator : SY/MD
Sample : N1559-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M51

Quant Time: Feb 17 04:52:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 16 13:52:04 2022
Response via : Initial Calibration

