

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021720\
 Data File : VU036807.D
 Acq On : 17 Feb 2020 14:05
 Operator : JC/MD
 Sample : L1537-14DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR4DL

Quant Time: Feb 18 01:33:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Feb 18 00:46:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	114163	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	104886	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	42384	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	40061	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	1.93	69	40030	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.59	63	55952	3.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.20%
20) 2-Butanone-d5	4.67	46	112549	54.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.34%
24) Chloroform-d	5.11	84	75042	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.74	65	43145	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.76	84	162474	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.73	67	53719	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.20%
41) Toluene-d8	7.92	98	144053	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	8.21	79	18861	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.66	63	93550	52.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.14%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	41018	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	43710	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	6275	0.536 ug/L	92
47) Tetrachloroethene	8.58	164	45867	8.183 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021720\
Data File : VU036807.D
Acq On : 17 Feb 2020 14:05
Operator : JC/MD
Sample : L1537-14DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR4DL

Quant Time: Feb 18 01:33:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Feb 18 00:46:02 2020
Response via : Initial Calibration

