

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021720\
 Data File : VU036820.D
 Acq On : 17 Feb 2020 19:28
 Operator : JC/MD
 Sample : L1540-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBD62

Quant Time: Feb 18 03:06:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Feb 18 00:46:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	112280	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	104521	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	40456	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	35905	3.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.40%
7) Chloroethane-d5	1.93	69	36440	4.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	2.59	63	50332	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	4.67	46	104669	51.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.38%
24) Chloroform-d	5.11	84	66093	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
26) 1,2-Dichloroethane-d4	5.74	65	40412	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.76	84	148063	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.73	67	49695	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.92	98	132598	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloropropene-	8.20	79	16572	4.23	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.60%
46) 2-Hexanone-d5	8.66	63	88148	49.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.48%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	38200	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	12.21	152	40132	4.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021720\
Data File : VU036820.D
Acq On : 17 Feb 2020 19:28
Operator : JC/MD
Sample : L1540-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBD62

Quant Time: Feb 18 03:06:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Feb 18 00:46:02 2020
Response via : Initial Calibration

