

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021721\
 Data File : VU042329.D
 Acq On : 17 Feb 2021 17:08
 Operator : SY/MD
 Sample : VU0217WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK22

Quant Time: Feb 18 03:49:31 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 18 03:48:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	157496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	229199	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	205084	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	99884	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	112347	47.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.08%	
35) Dibromofluoromethane	5.298	113	76323	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
50) Toluene-d8	7.906	98	280820	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	
62) 4-Bromofluorobenzene	10.639	95	102581	44.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021721\
Data File : VU042329.D
Acq On : 17 Feb 2021 17:08
Operator : SY/MD
Sample : VU0217WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK22

Quant Time: Feb 18 03:49:31 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 18 03:48:34 2021
Response via : Initial Calibration

