

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021722\
 Data File : VU047200.D
 Acq On : 17 Feb 2022 12:23
 Operator : SY/MD
 Sample : N1474-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Feb 18 00:56:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 18 00:54:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	365394	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	365877	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	200848	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	143240	43.164	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.320%
7) Chloroethane-d5	1.919	69	131704	53.538	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.080%
11) 1,1-Dichloroethene-d2	2.572	63	187022	34.008	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.020%
21) 2-Butanone-d5	4.629	46	295456	114.798	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.800%
24) Chloroform-d	5.067	84	272192	50.524	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.040%
26) 1,2-Dichloroethane-d4	5.707	65	182354	52.736	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.480%
32) Benzene-d6	5.732	84	557775	45.855	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.720%
36) 1,2-Dichloropropane-d6	6.694	67	184093	46.669	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.340%
41) Toluene-d8	7.903	98	535082	46.202	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.400%
43) trans-1,3-Dichloroprop...	8.182	79	91155	44.096	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.200%
47) 2-Hexanone-d5	8.633	63	214597	94.960	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.960%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	281843	49.001	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	224695	47.171	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021722\
Data File : VU047200.D
Acq On : 17 Feb 2022 12:23
Operator : SY/MD
Sample : N1474-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Feb 18 00:56:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 18 00:54:38 2022
Response via : Initial Calibration

