

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021821\
 Data File : VU042344.D
 Acq On : 18 Feb 2021 12:26
 Operator : SY/MD
 Sample : MDL03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Feb 19 03:08:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	128474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	201860	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	182799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	88182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	104465	53.62	ug/l	0.00
Spiked Amount 50.000			Recovery	= 107.24%		
35) Dibromofluoromethane	5.298	113	72894	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	= 102.32%		
50) Toluene-d8	7.906	98	257160	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	= 100.80%		
62) 4-Bromofluorobenzene	10.639	95	91431	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	= 90.92%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.388	85	631	0.41	ug/l	96
3) Chloromethane	1.523	50	893	0.56	ug/l	95
4) Vinyl Chloride	1.607	62	704	0.46	ug/l	97
5) Bromomethane	1.864	94	648	0.69	ug/l	99
6) Chloroethane	1.938	64	377	0.43	ug/l #	62
7) Trichlorofluoromethane	2.144	101	1025	0.44	ug/l	94
8) Diethyl Ether	2.385	74	299	0.40	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.585	101	469	0.39	ug/l	94
10) Methyl Iodide	2.729	142	950	0.55	ug/l #	85
11) Tert butyl alcohol	3.305	59	188	0.45	ug/l #	94
12) 1,1-Dichloroethene	2.588	96	477	0.40	ug/l	84
13) Acrolein	2.498	56	659	10.44	ug/l	89
14) Allyl chloride	2.929	41	1214	0.46	ug/l	94
15) Acrylonitrile	3.337	53	2181	2.17	ug/l	90
16) Acetone	2.652	43	3260	2.70	ug/l	96
17) Carbon Disulfide	2.800	76	1996	0.56	ug/l	99
18) Methyl Acetate	2.967	43	1058	0.48	ug/l #	72
19) Methyl tert-butyl Ether	3.375	73	2125	0.47	ug/l	96
20) Methylene Chloride	3.057	84	779	0.56	ug/l #	90
21) trans-1,2-Dichloroethene	3.366	96	601	0.47	ug/l #	62
22) Diisopropyl ether	3.999	45	2027	0.41	ug/l #	50
23) Vinyl Acetate	3.970	43	9449	2.00	ug/l	97
24) 1,1-Dichloroethane	3.874	63	1147	0.44	ug/l #	84
25) 2-Butanone	4.732	43	3679	2.29	ug/l	97
26) 2,2-Dichloropropane	4.678	77	1135	0.48	ug/l #	55
27) cis-1,2-Dichloroethene	4.678	96	551	0.37	ug/l	70
28) Bromochloromethane	4.986	49	665	0.50	ug/l #	62
29) Tetrahydrofuran	5.080	42	2366	2.37	ug/l	97
30) Chloroform	5.096	83	1162	0.44	ug/l	83
31) Cyclohexane	5.385	56	4347	1.70	ug/l #	33
32) 1,1,1-Trichloroethane	5.330	97	907	0.38	ug/l #	45
36) 1,1-Dichloropropene	5.536	75	868	0.43	ug/l #	73
38) Carbon Tetrachloride	5.533	117	787	0.37	ug/l #	84
39) Methylcyclohexane	6.774	83	1012	0.44	ug/l	91
40) Benzene	5.784	78	2251	0.40	ug/l	99
41) Methacrylonitrile	4.983	41	583	0.36	ug/l #	77

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021821\
 Data File : VU042344.D
 Acq On : 18 Feb 2021 12:26
 Operator : SY/MD
 Sample : MDL03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Feb 19 03:08:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 1,2-Dichloroethane	5.800	62	1352	0.55	ug/l #	79
43) Isopropyl Acetate	5.928	43	2142	0.47	ug/l #	73
44) Trichloroethene	6.555	130	512	0.34	ug/l	74
45) 1,2-Dichloropropane	6.803	63	617	0.40	ug/l	96
46) Dibromomethane	6.928	93	447	0.41	ug/l	91
47) Bromodichloromethane	7.115	83	1045	0.49	ug/l #	74
48) Methyl methacrylate	6.970	41	973	0.43	ug/l #	67
49) 1,4-Dioxane	7.018	88	145	3.35	ug/l #	12
51) 4-Methyl-2-Pentanone	7.806	43	6976	2.26	ug/l	96
52) Toluene	7.977	92	1662	0.48	ug/l	95
53) t-1,3-Dichloropropene	8.218	75	884	0.38	ug/l #	82
54) cis-1,3-Dichloropropene	7.616	75	905	0.36	ug/l #	87
55) 1,1,2-Trichloroethane	8.411	97	508	0.35	ug/l	88
56) Ethyl methacrylate	8.346	69	911	0.37	ug/l	93
57) 1,3-Dichloropropane	8.584	76	1088	0.43	ug/l	98
58) 2-Chloroethyl Vinyl ether	7.478	63	1131	21.04	ug/l	91
59) 2-Hexanone	8.703	43	5639	2.29	ug/l	91
60) Dibromochloromethane	8.819	129	594	0.35	ug/l	95
61) 1,2-Dibromoethane	8.935	107	637	0.39	ug/l	86
64) Tetrachloroethene	8.559	164	658	0.48	ug/l	92
65) Chlorobenzene	9.456	112	1749	0.48	ug/l	100
66) 1,1,1,2-Tetrachloroethane	9.542	131	490	0.35	ug/l #	66
67) Ethyl Benzene	9.578	91	2880	0.44	ug/l	98
68) m/p-Xylenes	9.700	106	1870	0.77	ug/l #	70
69) o-Xylene	10.105	106	948	0.40	ug/l	100
70) Styrene	10.121	104	1558	0.39	ug/l #	91
71) Bromoform	10.304	173	465	0.35	ug/l #	52
73) Isopropylbenzene	10.491	105	2835	0.48	ug/l	100
74) N-amyl acetate	10.330	43	1704	0.51	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	10.793	83	1123	0.50	ug/l #	88
76) 1,2,3-Trichloropropane	10.639	75	51552	23.64	ug/l #	37
77) Bromobenzene	10.790	156	811	0.52	ug/l	86
78) n-propylbenzene	10.912	91	3005	0.44	ug/l	96
79) 2-Chlorotoluene	10.996	91	1786	0.43	ug/l	91
80) 1,3,5-Trimethylbenzene	11.099	105	2317	0.46	ug/l	100
81) trans-1,4-Dichloro-2-b...	10.639	75	51552	64.28	ug/l #	7
82) 4-Chlorotoluene	11.108	91	2293	0.47	ug/l	99
83) tert-Butylbenzene	11.430	119	2271	0.46	ug/l	97
84) 1,2,4-Trimethylbenzene	11.475	105	2215	0.44	ug/l	99
85) sec-Butylbenzene	11.648	105	2390	0.41	ug/l	93
86) p-Isopropyltoluene	11.800	119	2413	0.45	ug/l	88
87) 1,3-Dichlorobenzene	11.755	146	1360	0.50	ug/l	99
88) 1,4-Dichlorobenzene	11.755	146	1360	0.49	ug/l	99
89) n-Butylbenzene	12.218	91	1953	0.42	ug/l	93
90) Hexachloroethane	12.488	117	284	0.30	ug/l	84
91) 1,2-Dichlorobenzene	12.224	146	1426	0.52	ug/l	90
92) 1,2-Dibromo-3-Chloropr...	13.009	75	281	0.46	ug/l	76
93) 1,2,4-Trichlorobenzene	13.848	180	795	2.08	ug/l	81
94) Hexachlorobutadiene	14.031	225	393	0.41	ug/l #	69
95) Naphthalene	14.095	128	2252	2.58	ug/l #	88
96) 1,2,3-Trichlorobenzene	14.340	180	906	2.14	ug/l	86

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021821\
Data File : VU042344.D
Acq On : 18 Feb 2021 12:26
Operator : SY/MD
Sample : MDL03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Feb 19 03:08:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Quant Time: Feb 19 03:08:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.382 min Scan# 1257
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:168 Resp: 128474
Ion Ratio Lower Upper
168 100
99 59.9 45.0 67.4



#2
Dichlorodifluoromethane
Concen: 0.41 ug/l
RT: 1.388 min Scan# 15
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 85 Resp: 631
Ion Ratio Lower Upper
85 100
87 30.6 16.5 49.5









#3
Chloromethane
Concen: 0.56 ug/l
RT: 1.523 min Scan# 57
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 50 Resp: 893
Ion Ratio Lower Upper
50 100
52 36.2 26.6 40.0



#4
Vinyl Chloride
Concen: 0.46 ug/l
RT: 1.607 min Scan# 83
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 62 Resp: 704
Ion Ratio Lower Upper
62 100
64 33.0 25.2 37.8









#5
Bromomethane
Concen: 0.69 ug/l
RT: 1.864 min Scan# 163
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 94 Resp: 648
Ion Ratio Lower Upper
94 100
96 94.9 76.6 115.0



#6
Chloroethane
Concen: 0.43 ug/l
RT: 1.938 min Scan# 186
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 64 Resp: 377
Ion Ratio Lower Upper
64 100
66 54.6 26.4 39.6#









#7
Trichlorofluoromethane
Concen: 0.44 ug/l
RT: 2.144 min Scan# 250
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:101 Resp: 1025
Ion Ratio Lower Upper
101 100
103 68.5 51.3 76.9



#8
Diethyl Ether
Concen: 0.40 ug/l
RT: 2.385 min Scan# 325
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 74 Resp: 299
Ion Ratio Lower Upper
74 100
45 110.4 56.9 170.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.39 ug/l
RT: 2.585 min Scan# 387
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:	101	Resp:	469
Ion	Ratio	Lower	Upper
101	100		
85	45.2	37.4	56.0
151	89.8	65.8	98.8



#10
Methyl Iodide
Concen: 0.55 ug/l
RT: 2.729 min Scan# 432
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion:	142	Resp:	950
Ion	Ratio	Lower	Upper
142	100		
127	41.5	37.8	56.6
141	0.0	11.8	17.6#








#11
Tert butyl alcohol
Concen: 0.45 ug/l
RT: 3.305 min Scan# 611
Delta R.T. 0.045 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampled :







Tgt Ion: 59 Resp: 188
Ion Ratio Lower Upper
59 100
57 0.0 1.6 2.4#



#12
1,1-Dichloroethene
Concen: 0.40 ug/l
RT: 2.588 min Scan# 388
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion: 96 Resp: 477
Ion Ratio Lower Upper
96 100
61 155.1 147.4 221.0
98 67.9 51.4 77.2








#13
Acrolein
Concen: 10.44 ug/l
RT: 2.498 min Scan# 360
Delta R.T. -0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion: 56 Resp: 659
Ion Ratio Lower Upper
56 100
55 80.6 57.3 85.9



#14
Allyl chloride
Concen: 0.46 ug/l
RT: 2.929 min Scan# 494
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 41 Resp: 1214
Ion Ratio Lower Upper
41 100
39 64.9 54.9 82.3
76 22.7 22.5 33.7









#15
Acrylonitrile
Concen: 2.17 ug/l
RT: 3.337 min Scan# 621
Delta R.T. 0.010 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 53 Resp: 2181
Ion Ratio Lower Upper
53 100
52 75.8 68.2 102.2
51 31.2 29.2 43.8



#16
Acetone
Concen: 2.70 ug/l
RT: 2.652 min Scan# 408
Delta R.T. 0.006 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion: 43 Resp: 3260
Ion Ratio Lower Upper
43 100
58 29.0 21.5 32.3








#17
Carbon Disulfide
Concen: 0.56 ug/l
RT: 2.800 min Scan# 454
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Instrument :
MSVOA_U
ClientSampled :





Tgt Ion: 76 Resp: 1996
Ion Ratio Lower Upper
76 100
78 9.7 7.5 11.3




#18
Methyl Acetate
Concen: 0.48 ug/l
RT: 2.967 min Scan# 506
Delta R.T. 0.010 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26



Tgt Ion: 43 Resp: 1058
Ion Ratio Lower Upper
43 100
74 7.9 16.6 25.0#









#19

Methyl tert-butyl Ether
Concen: 0.47 ug/l
RT: 3.375 min Scan# 633
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampled :




Tgt Ion: 73 Resp: 2125
Ion Ratio Lower Upper
73 100
57 25.5 18.8 28.2




#20

Methylene Chloride
Concen: 0.56 ug/l
RT: 3.057 min Scan# 534
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 84 Resp: 779
Ion Ratio Lower Upper
84 100
49 145.0 112.6 169.0
51 49.4 33.7 50.5
86 83.4 51.7 77.5#









#21
trans-1,2-Dichloroethene
Concen: 0.47 ug/l
RT: 3.366 min Scan# 630
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 96 Resp: 601
Ion Ratio Lower Upper
96 100
61 96.5 122.4 183.6#
98 47.2 51.9 77.9#



#22
Diisopropyl ether
Concen: 0.41 ug/l
RT: 3.999 min Scan# 827
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 45 Resp: 2027
Ion Ratio Lower Upper
45 100
43 141.2 73.0 109.6#
87 0.0 19.5 29.3#
59 0.0 8.6 12.8#









#23
 Vinyl Acetate
 Concen: 2.00 ug/l
 RT: 3.970 min Scan# 818
 Delta R.T. 0.003 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion: 43 Resp: 9449
 Ion Ratio Lower Upper
 43 100
 86 7.3 6.7 10.1



#24
 1,1-Dichloroethane
 Concen: 0.44 ug/l
 RT: 3.874 min Scan# 788
 Delta R.T. -0.006 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Tgt Ion: 63 Resp: 1147
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.1 9.3#
 100 0.0 2.1 6.2#





#25

2-Butanone

Concen: 2.29 ug/l

RT: 4.732 min Scan# 1055

Delta R.T. 0.016 min

Lab File: VU042344.D





Acq: 18 Feb 2021 12:26

Tgt Ion: 43 Resp: 3679

Ion Ratio Lower Upper

43 100

72 20.2 17.4 26.0



#26

2,2-Dichloropropane

Concen: 0.48 ug/l

RT: 4.678 min Scan# 1038

Delta R.T. 0.003 min

Lab File: VU042344.D




Acq: 18 Feb 2021 12:26


Tgt Ion: 77 Resp: 1135

Ion Ratio Lower Upper

77 100

97 0.0 10.6 31.8#









#27
 cis-1,2-Dichloroethene
 Concen: 0.37 ug/l
 RT: 4.678 min Scan# 1038
 Delta R.T. 0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion:	96	Resp:	551
Ion	Ratio	Lower	Upper
96	100		
61	210.0	0.0	322.2
98	53.7	0.0	128.6



#28
 Bromochloromethane
 Concen: 0.50 ug/l
 RT: 4.986 min Scan# 1134
 Delta R.T. 0.003 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Tgt Ion:	49	Resp:	665
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.4
130	39.2	57.1	85.7#





#29

Tetrahydrofuran

Concen: 2.37 ug/l

RT: 5.080 min Scan# 1163





Delta R.T. 0.016 min

Lab File: VU042344.D

Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampled :

Tgt Ion:	42	Resp:	2366
Ion	Ratio	Lower	Upper
42	100		
72	41.0	31.5	47.3
71	34.3	29.5	44.3



#30

Chloroform

Concen: 0.44 ug/l




RT: 5.096 min Scan# 1168


Delta R.T. -0.003 min

Lab File: VU042344.D

Acq: 18 Feb 2021 12:26

Tgt Ion:	83	Resp:	1162
Ion	Ratio	Lower	Upper
83	100		
85	54.6	54.6	81.8









#31
Cyclohexane
Concen: 1.70 ug/l
RT: 5.385 min Scan# 1258
Delta R.T. -0.013 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 56 Resp: 4347
Ion Ratio Lower Upper
56 100
69 137.3 24.9 37.3#
84 95.3 63.7 95.5



#32
1,1,1-Trichloroethane
Concen: 0.38 ug/l
RT: 5.330 min Scan# 1241
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 97 Resp: 907
Ion Ratio Lower Upper
97 100
99 0.0 51.4 77.0#
61 59.5 37.4 56.2#









#33
 1,2-Dichloroethane-d4
 Concen: 53.62 ug/l
 RT: 5.710 min Scan# 1359
 Delta R.T. 0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Instrument :
 MSVOA_U
 ClientSampled :

Tgt Ion: 65 Resp: 104465




Ion	Ratio	Lower	Upper
65	100		
67	49.6	0.0	99.8




#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.256 min Scan# 1529
 Delta R.T. -0.003 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Tgt Ion: 114 Resp: 201860

Ion	Ratio	Lower	Upper
114	100		
63	23.2	0.0	44.2
88	16.1	0.0	33.4









#35
Dibromofluoromethane
Concen: 51.16 ug/l
RT: 5.298 min Scan# 1231
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:	113	Resp:	72894
Ion	Ratio	Lower	Upper
113	100		
111	102.3	81.8	122.8
192	21.2	17.2	25.8



#36
1,1-Dichloropropene
Concen: 0.43 ug/l
RT: 5.536 min Scan# 1305
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:	75	Resp:	868
Ion	Ratio	Lower	Upper
75	100		
110	29.7	16.8	50.3
77	4.6	24.8	37.2#









#38
Carbon Tetrachloride
Concen: 0.37 ug/l
RT: 5.533 min Scan# 1304
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:117 Resp: 787
Ion Ratio Lower Upper
117 100
119 78.3 75.5 113.3
121 22.6 24.7 37.1#



#39
Methylcyclohexane
Concen: 0.44 ug/l
RT: 6.774 min Scan# 1690
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 83 Resp: 1012
Ion Ratio Lower Upper
83 100
55 75.2 66.8 100.2
98 51.6 36.9 55.3









#40
Benzene
Concen: 0.40 ug/l
RT: 5.784 min Scan# 1382
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 78 Resp: 2251
Ion Ratio Lower Upper
78 100
77 24.4 19.3 28.9



#41
Methacrylonitrile
Concen: 0.36 ug/l
RT: 4.983 min Scan# 1133
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 41 Resp: 583
Ion Ratio Lower Upper
41 100
39 64.8 45.1 67.7
67 47.0 50.2 75.2#
52 0.0 22.1 33.1#









#42
1,2-Dichloroethane
Concen: 0.55 ug/l
RT: 5.800 min Scan# 1387
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 62 Resp: 1352
Ion Ratio Lower Upper
62 100
98 0.0 0.0 14.8



#43
Isopropyl Acetate
Concen: 0.47 ug/l
RT: 5.928 min Scan# 1427
Delta R.T. 0.010 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 43 Resp: 2142
Ion Ratio Lower Upper
43 100
61 4.8 14.4 21.6#
87 3.0 8.8 13.2#









#44
Trichloroethene
Concen: 0.34 ug/l
RT: 6.555 min Scan# 1622
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:130 Resp: 512
Ion Ratio Lower Upper
130 100
95 124.4 0.0 197.8



#45
1,2-Dichloropropane
Concen: 0.40 ug/l
RT: 6.803 min Scan# 1699
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 63 Resp: 617
Ion Ratio Lower Upper
63 100
65 34.0 25.4 38.2





#46

Dibromomethane

Concen: 0.41 ug/l

RT: 6.928 min Scan# 1738





Delta R.T. 0.000 min

Lab File: VU042344.D

Acq: 18 Feb 2021 12:26

Tgt Ion: 93 Resp: 447

Ion	Ratio	Lower	Upper
93	100		
95	81.2	63.6	95.4
174	118.8	83.4	125.2



#47

Bromodichloromethane

Concen: 0.49 ug/l

RT: 7.115 min Scan# 1796




Delta R.T. 0.000 min


Lab File: VU042344.D

Acq: 18 Feb 2021 12:26

Tgt Ion: 83 Resp: 1045

Ion	Ratio	Lower	Upper
83	100		
85	43.9	51.3	76.9#
127	0.0	6.9	10.3#





#48

Methyl methacrylate

Concen: 0.43 ug/l

RT: 6.970 min Scan# 1751

Delta R.T. 0.003 min

Lab File: VU042344.D

Acq: 18 Feb 2021 12:26





Tgt Ion: 41 Resp: 973

Ion Ratio Lower Upper

41 100

69 50.7 57.0 85.6#

39 27.7 48.4 72.6#



#49

1,4-Dioxane

Concen: 3.35 ug/l

RT: 7.018 min Scan# 1766

Delta R.T. 0.003 min

Lab File: VU042344.D

Acq: 18 Feb 2021 12:26




Tgt Ion: 88 Resp: 145


Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 0.0 60.4 90.6#









#50
Toluene-d8
Concen: 50.40 ug/l
RT: 7.906 min Scan# 2042
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 98 Resp: 257160
Ion Ratio Lower Upper
98 100
100 61.9 50.2 75.2



#51
4-Methyl-2-Pentanone
Concen: 2.26 ug/l
RT: 7.806 min Scan# 2011
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 43 Resp: 6976
Ion Ratio Lower Upper
43 100
58 33.0 28.1 42.1









#52
Toluene
Concen: 0.48 ug/l
RT: 7.977 min Scan# 2064
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 92 Resp: 1662
Ion Ratio Lower Upper
92 100
91 168.1 140.3 210.5



#53
t-1,3-Dichloropropene
Concen: 0.38 ug/l
RT: 8.218 min Scan# 2139
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 75 Resp: 884
Ion Ratio Lower Upper
75 100
77 42.1 25.6 38.4#









#54
 cis-1,3-Dichloropropene
 Concen: 0.36 ug/l
 RT: 7.616 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion: 75 Resp: 905
 Ion Ratio Lower Upper
 75 100
 77 38.7 24.6 37.0#
 39 67.1 46.1 69.1



#55
 1,1,2-Trichloroethane
 Concen: 0.35 ug/l
 RT: 8.411 min Scan# 2199
 Delta R.T. 0.003 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Tgt Ion: 97 Resp: 508
 Ion Ratio Lower Upper
 97 100
 83 103.4 72.6 109.0
 85 63.9 46.6 69.8
 99 52.7 50.9 76.3









#56
Ethyl methacrylate
Concen: 0.37 ug/l
RT: 8.346 min Scan# 2179
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:	69	Resp:	911
Ion	Ratio	Lower	Upper
69	100		
41	86.1	62.8	94.2
39	47.1	36.2	54.2



#57
1,3-Dichloropropane
Concen: 0.43 ug/l
RT: 8.584 min Scan# 2253
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:	76	Resp:	1088
Ion	Ratio	Lower	Upper
76	100		
78	31.1	25.7	38.5









#58
2-Chloroethyl Vinyl ether
Concen: 21.04 ug/l
RT: 7.478 min Scan# 1909
Delta R.T. 0.006 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 63 Resp: 1131
Ion Ratio Lower Upper
63 100
106 20.8 20.3 30.5



#59
2-Hexanone
Concen: 2.29 ug/l
RT: 8.703 min Scan# 2290
Delta R.T. 0.006 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 43 Resp: 5639
Ion Ratio Lower Upper
43 100
58 41.9 24.0 72.0









#60
Dibromochloromethane
Concen: 0.35 ug/l
RT: 8.819 min Scan# 2326
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:129 Resp: 594
Ion Ratio Lower Upper
129 100
127 81.3 38.5 115.3



#61
1,2-Dibromoethane
Concen: 0.39 ug/l
RT: 8.935 min Scan# 2362
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:107 Resp: 637
Ion Ratio Lower Upper
107 100
109 81.8 75.9 113.9









#62
4-Bromofluorobenzene
Concen: 45.46 ug/l
RT: 10.639 min Scan# 2892
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 95 Resp: 91431
Ion Ratio Lower Upper
95 100
174 86.7 0.0 172.2
176 84.8 0.0 168.8



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 9.423 min Scan# 2514
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion: 117 Resp: 182799
Ion Ratio Lower Upper
117 100
82 57.5 45.8 68.6
119 31.6 25.8 38.8








#64
Tetrachloroethene
Concen: 0.48 ug/l
RT: 8.559 min Scan# 2245
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampleId :



Tgt Ion:164 Resp: 658




Ion	Ratio	Lower	Upper
164	100		
166	116.1	105.3	157.9
129	86.4	75.0	112.6
131	88.0	71.0	106.6




#65
Chlorobenzene
Concen: 0.48 ug/l
RT: 9.456 min Scan# 2524
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:112 Resp: 1749

Ion	Ratio	Lower	Upper
112	100		
114	31.3	25.0	37.4









#66
 1,1,1,2-Tetrachloroethane
 Concen: 0.35 ug/l
 RT: 9.542 min Scan# 2551
 Delta R.T. 0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26




Instrument :
 MSVOA_U
 ClientSampled :


Tgt Ion:131 Resp: 490
 Ion Ratio Lower Upper
 131 100
 133 92.9 47.6 142.9
 119 0.0 31.9 95.5#



#67
 Ethyl Benzene
 Concen: 0.44 ug/l
 RT: 9.578 min Scan# 2562
 Delta R.T. 0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Tgt Ion: 91 Resp: 2880
 Ion Ratio Lower Upper
 91 100
 106 31.4 24.1 36.1









#68
m/p-Xylenes
Concen: 0.77 ug/l
RT: 9.700 min Scan# 2600
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:106 Resp: 1870
Ion Ratio Lower Upper
106 100
91 259.7 169.9 254.9#



#69
o-Xylene
Concen: 0.40 ug/l
RT: 10.105 min Scan# 2726
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:106 Resp: 948
Ion Ratio Lower Upper
106 100
91 221.6 111.2 333.6









#70
Styrene
Concen: 0.39 ug/l
RT: 10.121 min Scan# 2731
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:104 Resp: 1558
Ion Ratio Lower Upper
104 100
78 54.4 44.8 67.2
103 45.0 45.7 68.5#



#71
Bromoform
Concen: 0.35 ug/l
RT: 10.304 min Scan# 2788
Delta R.T. 0.006 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:173 Resp: 465
Ion Ratio Lower Upper
173 100
175 15.1 23.8 71.2#
254 0.0 0.0 0.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 11.819 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:152 Resp: 88182
Ion Ratio Lower Upper
152 100
115 58.6 41.5 124.6
150 156.2 0.0 342.4



#73
Isopropylbenzene
Concen: 0.48 ug/l
RT: 10.491 min Scan# 2846
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:105 Resp: 2835
Ion Ratio Lower Upper
105 100
120 26.3 13.1 39.3









#74
N-amy1 acetate
Concen: 0.51 ug/l
RT: 10.330 min Scan# 2796
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:	43	Resp:	1704
Ion	Ratio	Lower	Upper
43	100		
70	29.2	28.3	42.5
55	20.2	17.2	25.8
61	11.9	17.3	25.9#



#75
1,1,2,2-Tetrachloroethane
Concen: 0.50 ug/l
RT: 10.793 min Scan# 2940
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:	83	Resp:	1123
Ion	Ratio	Lower	Upper
83	100		
131	3.6	5.1	15.2#
85	54.6	31.7	95.1









#76
1,2,3-Trichloropropane
Concen: 23.64 ug/l
RT: 10.639 min Scan# 2892
Delta R.T. -0.196 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 75 Resp: 51552
Ion Ratio Lower Upper
75 100
77 1.1 20.2 60.5#



#77
Bromobenzene
Concen: 0.52 ug/l
RT: 10.790 min Scan# 2939
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:156 Resp: 811
Ion Ratio Lower Upper
156 100
77 175.2 79.3 238.1
158 81.0 48.3 144.9









#78
n-propylbenzene
Concen: 0.44 ug/l
RT: 10.912 min Scan# 2977
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 91 Resp: 3005
Ion Ratio Lower Upper
91 100
120 24.2 11.2 33.6



#79
2-Chlorotoluene
Concen: 0.43 ug/l
RT: 10.996 min Scan# 3003
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 91 Resp: 1786
Ion Ratio Lower Upper
91 100
126 38.5 16.7 50.0









#80
1,3,5-Trimethylbenzene
Concen: 0.46 ug/l
RT: 11.099 min Scan# 3035
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion:105 Resp: 2317
Ion Ratio Lower Upper
105 100
120 47.5 23.8 71.4



#81
trans-1,4-Dichloro-2-butene
Concen: 64.28 ug/l
RT: 10.639 min Scan# 2892
Delta R.T. 0.084 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 75 Resp: 51552
Ion Ratio Lower Upper
75 100
53 0.0 85.6 128.4#
89 0.0 35.7 53.5#









#82
4-Chlorotoluene
Concen: 0.47 ug/l
RT: 11.108 min Scan# 3038
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 91 Resp: 2293
Ion Ratio Lower Upper
91 100
126 29.9 14.6 44.0



#83
tert-Butylbenzene
Concen: 0.46 ug/l
RT: 11.430 min Scan# 3138
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion: 119 Resp: 2271
Ion Ratio Lower Upper
119 100
91 62.5 30.3 90.9
134 21.3 11.4 34.1









#84
1,2,4-Trimethylbenzene
Concen: 0.44 ug/l
RT: 11.475 min Scan# 3152
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:105 Resp: 2215
Ion Ratio Lower Upper
105 100
120 45.0 22.3 66.8



#85
sec-Butylbenzene
Concen: 0.41 ug/l
RT: 11.648 min Scan# 3206
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:105 Resp: 2390
Ion Ratio Lower Upper
105 100
134 17.0 10.0 30.0









#86
p-Isopropyltoluene
Concen: 0.45 ug/l
RT: 11.800 min Scan# 3253
Delta R.T. -0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:119 Resp: 2413
Ion Ratio Lower Upper
119 100
134 24.3 13.2 39.6
91 35.5 12.8 38.3



#87
1,3-Dichlorobenzene
Concen: 0.50 ug/l
RT: 11.755 min Scan# 3239
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion:146 Resp: 1360
Ion Ratio Lower Upper
146 100
111 39.2 19.9 59.6
148 64.3 32.0 96.2








#88
1,4-Dichlorobenzene
Concen: 0.49 ug/l
RT: 11.755 min Scan# 3239
Delta R.T. -0.090 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion:146 Resp: 1360
Ion Ratio Lower Upper
146 100
111 39.2 19.7 59.0
148 64.3 32.8 98.3



#89
n-Butylbenzene
Concen: 0.42 ug/l
RT: 12.218 min Scan# 3383
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion: 91 Resp: 1953
Ion Ratio Lower Upper
91 100
92 55.5 26.2 78.6
134 19.7 12.7 38.1








#90
Hexachloroethane
Concen: 0.30 ug/l
RT: 12.488 min Scan# 3467
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Instrument :
MSVOA_U
ClientSampled :





Tgt Ion:117 Resp: 284
Ion Ratio Lower Upper
117 100
201 101.8 43.5 130.7




#91
1,2-Dichlorobenzene
Concen: 0.52 ug/l
RT: 12.224 min Scan# 3385
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26



Tgt Ion:146 Resp: 1426
Ion Ratio Lower Upper
146 100
111 29.9 20.8 62.4
148 60.2 31.9 95.7









#92
1,2-Dibromo-3-Chloropropane
Concen: 0.46 ug/l
RT: 13.009 min Scan# 3629
Delta R.T. 0.003 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampled :


Tgt Ion: 75 Resp: 281
Ion Ratio Lower Upper
75 100
155 82.2 41.7 125.1
157 65.1 53.9 161.7



#93
1,2,4-Trichlorobenzene
Concen: 2.08 ug/l
RT: 13.848 min Scan# 3890
Delta R.T. -0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26

Tgt Ion:180 Resp: 795
Ion Ratio Lower Upper
180 100
182 109.9 46.9 140.8
145 15.1 14.4 43.4









#94
Hexachlorobutadiene
Concen: 0.41 ug/l
RT: 14.031 min Scan# 3947
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion:225 Resp: 393
Ion Ratio Lower Upper
225 100
223 30.8 32.3 96.9#
227 48.3 31.8 95.4



#95
Naphthalene
Concen: 2.58 ug/l
RT: 14.095 min Scan# 3967
Delta R.T. 0.000 min
Lab File: VU042344.D
Acq: 18 Feb 2021 12:26


Tgt Ion:128 Resp: 2252
Ion Ratio Lower Upper
128 100
127 5.9 10.8 16.2#
129 9.3 8.6 13.0





#96
 1,2,3-Trichlorobenzene
 Concen: 2.14 ug/l
 RT: 14.340 min Scan# 4043
 Delta R.T. -0.000 min
 Lab File: VU042344.D
 Acq: 18 Feb 2021 12:26

Instrument :
 MSVOA_U
 ClientSampleId :



Tgt Ion:	180	Resp:	906
Ion	Ratio	Lower	Upper
180	100		
182	80.6	48.0	144.0
145	25.6	15.5	46.5

