

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU021821\  
 Data File : VU042339.D  
 Acq On : 18 Feb 2021 10:08  
 Operator : SY/MD  
 Sample : VU0218WBL03  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VU0218WBL03

Quant Time: Feb 19 03:07:08 2021  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\82U021721W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 19 02:55:52 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) Pentafluorobenzene       | 5.385  | 168  | 150874   | 50.00 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.260  | 114  | 224640   | 50.00 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 9.427  | 117  | 200679   | 50.00 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.822 | 152  | 95787    | 50.00 | ug/l   | 0.00     |
| System Monitoring Compounds |        |      |          |       |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.713  | 65   | 106959   | 46.75 | ug/l   | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 93.50% |          |
| 35) Dibromofluoromethane    | 5.301  | 113  | 75251    | 47.46 | ug/l   | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 94.92% |          |
| 50) Toluene-d8              | 7.906  | 98   | 274034   | 48.26 | ug/l   | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 96.52% |          |
| 62) 4-Bromofluorobenzene    | 10.642 | 95   | 101805   | 45.49 | ug/l   | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 90.98% |          |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU021821\  
Data File : VU042339.D  
Acq On : 18 Feb 2021 10:08  
Operator : SY/MD  
Sample : VU0218WBL03  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0218WBL03

Quant Time: Feb 19 03:07:08 2021  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\82U021721W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 19 02:55:52 2021  
Response via : Initial Calibration

