

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021822\
 Data File : VU047214.D
 Acq On : 17 Feb 2022 19:10
 Operator : SY/MD
 Sample : N1559-11DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M85DL

Quant Time: Feb 18 03:35:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Feb 18 03:32:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	158455	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	163714	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	90651	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	58341	4.359	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.200%	
7) Chloroethane-d5	1.919	69	51064	4.815	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.572	65	23090	4.172	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	83.400%	
20) 2-Butanone-d5	4.639	46	169841	53.097	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	106.200%	
24) Chloroform-d	5.067	84	108823	4.895	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
26) 1,2-Dichloroethane-d4	5.707	65	58891	4.764	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.200%	
32) Benzene-d6	5.732	84	219887	4.807	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
36) 1,2-Dichloropropane-d6	6.694	67	71161	4.889	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.800%	
41) Toluene-d8	7.903	98	208768	4.665	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.400%	
43) trans-1,3-Dichloroprop...	8.182	79	27729	4.203	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.000%	
46) 2-Hexanone-d5	8.636	63	147817	42.547	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	85.100%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	56174	4.246	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.000%	
66) 1,2-Dichlorobenzene-d4	12.195	152	82445	5.069	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.400%	
Target Compounds						
					Qvalue	
42) Toluene	7.973	91	244153	4.715	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021822\
Data File : VU047214.D
Acq On : 17 Feb 2022 19:10
Operator : SY/MD
Sample : N1559-11DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M85DL

Quant Time: Feb 18 03:35:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR021622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Feb 18 03:32:37 2022
Response via : Initial Calibration

