

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU021919\
 Data File : VU029557.D
 Acq On : 19 Feb 2019 16:04
 Operator : JC/SP
 Sample : K1488-20DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ESXF7DL

Quant Time: Feb 25 03:06:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Feb 20 00:07:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	211336	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	204133	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	109803	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	56810	4.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.00%
7) Chloroethane-d5	1.68	69	53635	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.60%
11) 1,1-Dichloroethene-d2	2.27	63	91532	2.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.40%#
20) 2-Butanone-d5	4.19	46	153497	51.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.98%
24) Chloroform-d	4.65	84	101264	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
26) 1,2-Dichloroethane-d4	5.31	65	58108	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.34	84	223048	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
36) 1,2-Dichloropropane-d6	6.33	67	70370	4.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.00%
41) Toluene-d8	7.56	98	212886	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.85	79	26578	4.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.00%
46) 2-Hexanone-d5	8.31	63	152850	51.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.48%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	60254	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	97997	4.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds

					Qvalue
5) Vinyl chloride	1.40	62	21631	1.063 ug/L #	73
18) trans-1,2-Dichloroethene	2.98	96	11065	0.597 ug/L	99
22) cis-1,2-Dichloroethene	4.22	96	129070	7.566 ug/L	92
34) Trichloroethene	6.18	95	24382	1.443 ug/L	95
47) Tetrachloroethene	8.22	164	46173	2.984 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU021919\
Data File : VU029557.D
Acq On : 19 Feb 2019 16:04
Operator : JC/SP
Sample : K1488-20DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ESXF7DL

Quant Time: Feb 25 03:06:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Feb 20 00:07:58 2019
Response via : Initial Calibration

