

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU021920\
 Data File : VU036852.D
 Acq On : 19 Feb 2020 17:16
 Operator : JC/MD
 Sample : L1580-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBDJ9

Quant Time: Feb 20 06:03:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 20 04:46:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	136366	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	127815	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	62500	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	47833	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	1.93	69	43623	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.60%
11) 1,1-Dichloroethene-d2	2.59	63	68205	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	4.67	46	118127	48.03	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.06%
24) Chloroform-d	5.11	84	78835	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.74	65	47764	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
32) Benzene-d6	5.76	84	184225	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.73	67	58024	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.93	98	167992	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
43) trans-1,3-Dichloropropene-	8.20	79	20107	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
46) 2-Hexanone-d5	8.66	63	105342	48.12	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.24%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	44648	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	54768	4.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.66	43	4310	1.371	ug/L	98
42) Toluene	8.00	91	38959	0.920	ug/L	99
52) Ethylbenzene	9.59	91	1551681	32.812	ug/L	100
53) m,p-Xylene	9.71	106	1210628	69.562	ug/L	94
54) o-Xylene	10.12	106	531785	31.116	ug/L	96
55) Styrene	10.13	104	92976	3.227	ug/L #	27
56) Isopropylbenzene	10.50	105	199061	4.352	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU021920\
Data File : VU036852.D
Acq On : 19 Feb 2020 17:16
Operator : JC/MD
Sample : L1580-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBDJ9

Quant Time: Feb 20 06:03:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 20 04:46:21 2020
Response via : Initial Calibration

