

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022024\
 Data File : VU057895.D
 Acq On : 20 Feb 2024 12:17
 Operator : MD/SY
 Sample : P1496-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHA01

Quant Time: Feb 21 00:13:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 21 00:11:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	280569	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	264710	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	124717	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	94957	46.748	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.500%
7) Chloroethane-d5	1.888	69	72586	48.347	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.700%
11) 1,1-Dichloroethene-d2	2.554	63	128323	33.898	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.800%
21) 2-Butanone-d5	4.608	46	93443	121.075	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	121.080%
24) Chloroform-d	5.052	84	158554	46.033	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.060%
26) 1,2-Dichloroethane-d4	5.689	65	106070	49.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.260%
32) Benzene-d6	5.714	84	337888	50.934	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.860%
36) 1,2-Dichloropropane-d6	6.679	67	109517	51.127	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.260%
41) Toluene-d8	7.888	98	297262	49.502	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.000%
43) trans-1,3-Dichloroprop...	8.174	79	48352	47.401	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.800%
47) 2-Hexanone-d5	8.627	63	73619	89.400	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	132831	44.939	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.880%
66) 1,2-Dichlorobenzene-d4	12.187	152	101256	50.141	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.280%
Target Compounds						
13) Acetone	2.615	43	8423	5.894	ug/L	Qvalue 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022024\
Data File : VU057895.D
Acq On : 20 Feb 2024 12:17
Operator : MD/SY
Sample : P1496-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHA01

Quant Time: Feb 21 00:13:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 21 00:11:26 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 5.894 ug/L
 RT: 2.615 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VU057895.D
 Acq: 20 Feb 2024 12:17

Instrument :
 MSVOA_U
 ClientSampleId :
 BHA01

Tgt Ion: 43 Resp: 8423
 Ion Ratio Lower Upper
 43 100
 58 26.5 0.0 64.2

