

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022024\
 Data File : VU057908.D
 Acq On : 20 Feb 2024 17:31
 Operator : MD/SY
 Sample : P1496-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHA18

Quant Time: Feb 21 00:16:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 21 00:11:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	291755	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	275718	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	125340	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	89206	42.233	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.460%
7) Chloroethane-d5	1.907	69	71105	45.545	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.100%
11) 1,1-Dichloroethene-d2	2.560	63	124822	31.709	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.420%
21) 2-Butanone-d5	4.611	46	82818	103.194	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.190%
24) Chloroform-d	5.052	84	163931	45.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.540%
26) 1,2-Dichloroethane-d4	5.692	65	106907	47.615	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.240%
32) Benzene-d6	5.717	84	330772	47.871	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.740%
36) 1,2-Dichloropropane-d6	6.682	67	108553	48.654	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.300%
41) Toluene-d8	7.891	98	293100	46.861	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.720%
43) trans-1,3-Dichloroprop...	8.174	79	47599	44.800	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.600%
47) 2-Hexanone-d5	8.624	63	70406	82.084	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.080%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	129031	41.911	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.820%
66) 1,2-Dichlorobenzene-d4	12.187	152	96746	47.669	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.340%
Target Compounds						
					Qvalue	
13) Acetone	2.615	43	14634	9.848	ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022024\
Data File : VU057908.D
Acq On : 20 Feb 2024 17:31
Operator : MD/SY
Sample : P1496-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHA18

Quant Time: Feb 21 00:16:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 21 00:11:26 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 9.848 ug/L
 RT: 2.615 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VU057908.D
 Acq: 20 Feb 2024 17:31

Instrument :
 MSVOA_U
 ClientSampleId :
 BHA18

Tgt Ion: 43 Resp: 14634
 Ion Ratio Lower Upper
 43 100
 58 32.2 0.0 64.2

