

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022119\
 Data File : VU029593.D
 Acq On : 21 Feb 2019 13:10
 Operator : JC/SP
 Sample : VU0221WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK58

Quant Time: Feb 22 03:21:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 22 03:07:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	240506	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	228775	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	118087	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	82597	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.68	69	76224	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.20%
11) 1,1-Dichloroethene-d2	2.27	63	132395	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.40%
20) 2-Butanone-d5	4.19	46	192745	54.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.56%
24) Chloroform-d	4.65	84	148223	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
26) 1,2-Dichloroethane-d4	5.31	65	79276	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.33	84	317932	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
36) 1,2-Dichloropropane-d6	6.33	67	95516	5.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.80%
41) Toluene-d8	7.56	98	298160	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.60%
43) trans-1,3-Dichloropropene-	7.85	79	36048	5.14	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.80%
46) 2-Hexanone-d5	8.31	63	193615	56.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.62%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	79090	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	125636	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022119\
Data File : VU029593.D
Acq On : 21 Feb 2019 13:10
Operator : JC/SP
Sample : VU0221WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK58

Quant Time: Feb 22 03:21:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 22 03:07:37 2019
Response via : Initial Calibration

