

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022120\
 Data File : VU036874.D
 Acq On : 21 Feb 2020 11:33
 Operator : JC/MD
 Sample : VU0221WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK20

Quant Time: Feb 22 00:11:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Feb 22 00:05:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	131415	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	125380	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	50213	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	46394	4.10	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.00%
7) Chloroethane-d5	1.93	69	44725	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.59	63	63687	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	4.69	46	131336	55.42	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.84%
24) Chloroform-d	5.11	84	83296	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.74	65	48883	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
32) Benzene-d6	5.76	84	179927	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.73	67	58128	4.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.20%
41) Toluene-d8	7.93	98	161005	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	8.21	79	21136	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.66	63	107987	50.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.56%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	45734	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	12.21	152	50307	5.02	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.40%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	1462	0.109 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022120\
Data File : VU036874.D
Acq On : 21 Feb 2020 11:33
Operator : JC/MD
Sample : VU0221WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK20

Quant Time: Feb 22 00:11:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Feb 22 00:05:53 2020
Response via : Initial Calibration

