

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022120\
 Data File : VU036875.D
 Acq On : 21 Feb 2020 12:02
 Operator : JC/MD
 Sample : L1580-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Feb 22 00:17:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Feb 22 00:05:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	130866	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	123165	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	49513	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	44773	3.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.60%
7) Chloroethane-d5	1.93	69	42497	4.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.00%
11) 1,1-Dichloroethene-d2	2.59	63	61516	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	4.69	46	125371	53.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.24%
24) Chloroform-d	5.11	84	79930	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.74	65	46612	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	5.76	84	173617	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.73	67	56820	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.80%
41) Toluene-d8	7.92	98	154411	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
43) trans-1,3-Dichloropropene-	8.21	79	20271	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.67	63	100485	47.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.26%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	43902	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
64) 1,2-Dichlorobenzene-d4	12.21	152	47310	4.78	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.60%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	2161	0.161 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022120\
Data File : VU036875.D
Acq On : 21 Feb 2020 12:02
Operator : JC/MD
Sample : L1580-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Feb 22 00:17:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR021220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Feb 22 00:05:53 2020
Response via : Initial Calibration

