

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
 Data File : VU047227.D
 Acq On : 21 Feb 2022 12:52
 Operator : SY/MD
 Sample : N1205-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Feb 22 03:59:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	52343	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	18619	0.918	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	19667	0.931	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
Target Compounds						
						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.591	101	325253	27.056	ug/l	93
24) tert-Butyl Alcohol	3.289	59	25581	39.336	ug/l	# 1
25) Methyl tert-Butyl Ether	3.372	73	1434290	41.339	ug/l	100
31) Isopropyl Ether	4.002	45	260782	6.815	ug/l	99
32) Ethyl-t-butyl ether	4.510	59	1182878	30.411	ug/l	97
33) Tert-Amyl methyl ether	5.948	73	1605099	44.802	ug/l	98
73) n-propylbenzene	10.909	120	788696	42.324	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
Data File : VU047227.D
Acq On : 21 Feb 2022 12:52
Operator : SY/MD
Sample : N1205-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Feb 22 03:59:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

