

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
 Data File : VU047228.D
 Acq On : 21 Feb 2022 13:29
 Operator : SY/MD
 Sample : N1205-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-THM-WS

Quant Time: Feb 22 03:59:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	51777	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	19165	0.956	ug/l	0.00
Spiked Amount 1.000			Recovery	=	96.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	19846	0.950	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.000%	
Target Compounds						
27) Chloroform	5.095	83	1131063	47.101	ug/l	Qvalue 100
44) Bromodichloromethane	7.108	83	610794	33.932	ug/l	98
55) Dibromochloromethane	8.812	129	189627	14.363	ug/l	99
67) Bromoform	10.295	173	379851	45.402	ug/l	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
Data File : VU047228.D
Acq On : 21 Feb 2022 13:29
Operator : SY/MD
Sample : N1205-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-THM-WS

Quant Time: Feb 22 03:59:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

