

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
 Data File : VU047223.D
 Acq On : 21 Feb 2022 10:18
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 22 03:57:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.333	0.333	0.0	93	0.00
3 t	Chloromethane	0.342	0.509	-48.8#	139	0.00
4 Rt	Vinyl Chloride	0.338	0.331	2.1	92	0.00
5 T	Bromomethane	0.221	0.094	57.5#	43	0.00
6 T	Chloroethane	0.194	0.195	-0.5	95	0.00
7 T	Trichlorofluoromethane	0.398	0.431	-8.3	101	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.230	0.239	-3.9	97	0.00
9 Rt	1,1-Dichloroethene	0.236	0.239	-1.3	98	0.00
10 t	Iodomethane	0.223	0.106	52.5#	39	0.00
11 t	Allyl Chloride	0.368	0.333	9.5	85	0.00
12 t	Acrylonitrile	0.069	0.064	7.2	101	0.01
13 T	Acetone	0.040	0.053	-32.5#	140	0.03
14 T	Carbon Disulfide	0.800	0.766	4.3	90	0.00
15 RT	Methylene Chloride	0.265	0.282	-6.4	103	0.00
16 RT	trans-1,2-Dichloroethene	0.249	0.255	-2.4	96	0.00
17 t	1,1-Dichloroethane	0.457	0.453	0.9	93	0.00
18 T	2-Butanone	0.070	0.082	-17.1	116	0.02
19	Cyclohexane	0.423	0.409	3.3	90	0.00
20	Methylcyclohexane	0.467	0.473	-1.3	94	0.00
21 T	2,2-Dichloropropane	0.408	0.424	-3.9	96	0.00
22 RT	cis-1,2-Dichloroethene	0.287	0.286	0.3	95	0.00
23 t	Diethyl Ether	0.191	0.192	-0.5	95	0.00
24 t	tert-Butyl Alcohol	0.012	0.013	-8.3	96	0.10
25 t	Methyl tert-Butyl Ether	0.663	0.662	0.2	95	0.00
26 t	Bromochloromethane	0.129	0.118	8.5	85	0.00
27 t	Chloroform	0.464	0.462	0.4	96	0.00
28 RT	1,1,1-Trichloroethane	0.402	0.411	-2.2	96	0.00
29 T	1,1-Dichloropropene	0.368	0.373	-1.4	95	0.00
30 RT	Carbon Tetrachloride	0.353	0.360	-2.0	94	0.00
31 t	Isopropyl Ether	0.731	0.722	1.2	92	0.00
32	Ethyl-t-butyl ether	0.743	0.753	-1.3	94	0.00
33	Tert-Amyl methyl ether	0.684	0.697	-1.9	95	0.00
34 t	Propionitrile	0.019	0.025	-31.6#	135	0.02
35 RT	Benzene	1.030	1.043	-1.3	94	0.00
36 RT	1,2-Dichloroethane	0.312	0.318	-1.9	95	0.00
37 RT	Trichloroethene	0.285	0.298	-4.6	98	0.00
38 Rt	1,2-Dichloropropane	0.273	0.271	0.7	93	0.00
39 t	Methacrylonitrile	0.094	0.100	-6.4	104	0.00
40 t	Methyl acrylate	0.159	0.186	-17.0	115	0.00
41 t	Tetrahydrofuran	0.048	0.051	-6.2	116	0.01
42 t	1-Chlorobutane	0.494	0.493	0.2	93	0.00
43 T	Dibromomethane	0.147	0.152	-3.4	96	0.00
44 T	Bromodichloromethane	0.348	0.353	-1.4	93	0.00
45 T	4-Methyl-2-Pentanone	0.199	0.191	4.0	88	0.00
46 t	t-1,4-Dichloro-2-butene	0.087	0.080	8.0	104	0.00
47 t	Methyl methacrylate	0.154	0.154	0.0	96	0.00
48 t	Ethyl methacrylate	0.313	0.314	-0.3	90	0.00
49 Rt	Toluene	0.664	0.677	-2.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
 Data File : VU047223.D
 Acq On : 21 Feb 2022 10:18
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 22 03:57:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.376	0.382	-1.6	92	0.00
51 T	cis-1,3-Dichloropropene	0.422	0.426	-0.9	92	0.00
52 RT	1,1,2-Trichloroethane	0.210	0.214	-1.9	96	0.00
53 t	1,3-Dichloropropane	0.375	0.368	1.9	90	0.00
54 t	2-Hexanone	0.148	0.139	6.1	88	0.00
55 t	Dibromochloromethane	0.255	0.256	-0.4	90	0.00
56 T	1,2-Dibromoethane	0.212	0.217	-2.4	93	0.00
57 S	4-Bromofluorobenzene	0.387	0.374	3.4	94	0.00
58 RT	Tetrachloroethene	0.263	0.272	-3.4	96	0.00
59 Rt	Chlorobenzene	0.741	0.748	-0.9	94	0.00
60 T	1,1,1,2-Tetrachloroethane	0.255	0.264	-3.5	94	0.00
61 t	Pentachloroethane	0.193	0.211	-9.3	96	0.00
62 t	Hexachloroethane	0.209	0.194	7.2	82	0.00
63 Rt	Ethyl Benzene	1.302	1.302	0.0	92	0.00
64 RT	m/p-Xylenes	0.496	0.500	-0.8	92	0.00
65 RT	o-Xylene	0.481	0.482	-0.2	91	0.00
66 RT	Styrene	0.816	0.818	-0.2	91	0.00
67 t	Bromoform	0.162	0.152	6.2	84	0.00
68 S	1,2-Dichlorobenzene-d4	0.403	0.384	4.7	92	0.00
69 T	Isopropylbenzene	1.293	1.297	-0.3	92	0.00
70 T	1,1,2,2-Tetrachloroethane	0.284	0.288	-1.4	94	0.00
71 T	1,2,3-Trichloropropane	0.239	0.227	5.0	83	0.00
72 t	Bromobenzene	0.316	0.320	-1.3	93	0.00
73 t	n-propylbenzene	0.356	0.365	-2.5	93	0.00
74 t	2-Chlorotoluene	0.304	0.310	-2.0	93	0.00
75 t	1,3,5-Trimethylbenzene	1.061	1.076	-1.4	92	0.00
76 t	4-Chlorotoluene	0.319	0.329	-3.1	94	0.00
77 t	tert-Butylbenzene	1.064	1.093	-2.7	93	0.00
78 t	1,2,4-Trimethylbenzene	1.082	1.091	-0.8	92	0.00
79 t	sec-Butylbenzene	1.413	1.431	-1.3	91	0.00
80	Nitrobenzene	0.014	0.007	50.0#	46	0.00
81 t	p-Isopropyltoluene	1.196	1.216	-1.7	92	0.00
82 t	1,3-Dichlorobenzene	0.615	0.626	-1.8	93	0.00
83 Rt	1,4-Dichlorobenzene	0.619	0.626	-1.1	94	0.00
84 t	n-Butylbenzene	1.148	1.173	-2.2	92	0.00
85 Rt	1,2-Dichlorobenzene	0.592	0.595	-0.5	93	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.052	0.051	1.9	91	0.00
87 Rt	1,2,4-Trichlorobenzene	0.465	0.474	-1.9	93	0.00
88 t	Hexachlorobutadiene	0.238	0.249	-4.6	96	0.00
89 t	Naphthalene	0.952	0.923	3.0	90	0.00
90 t	1,2,3-Trichlorobenzene	0.417	0.429	-2.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0