

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
 Data File : VU047224.D
 Acq On : 21 Feb 2022 11:13
 Operator : SY/MD
 Sample : VU0221WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0221WBL01

Quant Time: Feb 22 03:57:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	53476	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.635	95	19282	0.931	ug/l	0.00
Spiked Amount 1.000			Recovery	=	93.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	19659	0.911	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.000%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022122\
Data File : VU047224.D
Acq On : 21 Feb 2022 11:13
Operator : SY/MD
Sample : VU0221WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0221WBL01

Quant Time: Feb 22 03:57:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.118 min Scan# 1486
Delta R.T. -0.000 min
Lab File: VU047224.D
Acq: 21 Feb 2022 11:13




Instrument :
MSVOA_U
ClientSampleId :
VU0221WBL01


Tgt Ion:	96	Resp:	53476
Ion	Ratio	Lower	Upper
96	100		
70	19.9	16.8	25.2
95	8.4	7.3	10.9
97	6.5	0.0	0.0#



#57
4-Bromofluorobenzene
Concen: 0.931 ug/l
RT: 10.635 min Scan# 2891
Delta R.T. -0.000 min
Lab File: VU047224.D
Acq: 21 Feb 2022 11:13

Tgt Ion:	95	Resp:	19282
Ion	Ratio	Lower	Upper
95	100		
174	84.9	56.6	84.8#
176	83.2	54.3	81.5#





#68
 1,2-Dichlorobenzene-d4
 Concen: 0.911 ug/l
 RT: 12.195 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: VU047224.D
 Acq: 21 Feb 2022 11:13

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0221WBL01

Tgt Ion:	152	Resp:	19659
Ion Ratio	Lower	Upper	
152	100		
115	62.6	0.0	251.0
150	161.3	0.0	613.8

