

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022219\  
 Data File : VU029629.D  
 Acq On : 22 Feb 2019 18:56  
 Operator : JC/SP  
 Sample : K1581-02  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 28 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 DB613

Quant Time: Feb 22 23:28:14 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM021519WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Fri Feb 22 23:08:46 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.88  | 114  | 462642   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.09  | 117  | 450932   | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 217233   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |         |
|--------------------------------|---------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.40    | 65    | 131920   | 38.91    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 77.82%  |
| 7) Chloroethane-d5             | 1.68    | 69    | 122087   | 40.51    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 81.02%  |
| 11) 1,1-Dichloroethene-d2      | 2.27    | 63    | 184432   | 27.83    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 55.66%# |
| 21) 2-Butanone-d5              | 4.17    | 46    | 215503   | 104.46   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 104.46% |
| 24) Chloroform-d               | 4.65    | 84    | 259166   | 47.13    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 94.26%  |
| 26) 1,2-Dichloroethane-d4      | 5.31    | 65    | 168581   | 48.43    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 96.86%  |
| 32) Benzene-d6                 | 5.34    | 84    | 533754   | 47.04    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 94.08%  |
| 36) 1,2-Dichloropropane-d6     | 6.33    | 67    | 170811   | 50.04    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 100.08% |
| 41) Toluene-d8                 | 7.56    | 98    | 500733   | 43.24    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 86.48%  |
| 43) trans-1,3-Dichloropropene- | 7.85    | 79    | 76774    | 42.40    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 84.80%  |
| 47) 2-Hexanone-d5              | 8.31    | 63    | 181684   | 95.57    | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 95.57%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.43   | 84    | 271127   | 47.00    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 94.00%  |
| 64) 1,2-Dichlorobenzene-d4     | 11.86   | 152   | 217790   | 46.52    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 93.04%  |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022219\  
Data File : VU029629.D  
Acq On : 22 Feb 2019 18:56  
Operator : JC/SP  
Sample : K1581-02  
Misc : 5.0mL/MSVOA U/WATER  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
DB613

Quant Time: Feb 22 23:28:14 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM021519WMA.M  
Quant Title : VOC Analysis  
QLast Update : Fri Feb 22 23:08:46 2019  
Response via : Initial Calibration

