

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
 Data File : VU042397.D
 Acq On : 22 Feb 2021 16:06
 Operator : SY/MD
 Sample : M1457-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GWBR1-70-85-021821

Quant Time: Feb 22 16:28:33 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	180193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	270177	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	241277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	111940	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	130346	47.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.40%	
35) Dibromofluoromethane	5.298	113	91867	48.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.34%	
50) Toluene-d8	7.906	98	329805	48.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.58%	
62) 4-Bromofluorobenzene	10.639	95	118167	43.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.80%	
Target Compounds						
						Qvalue
16) Acetone	2.642	43	14231	8.40	ug/l	98
21) trans-1,2-Dichloroethene	3.362	96	1768	0.99	ug/l #	76
27) cis-1,2-Dichloroethene	4.677	96	20788	10.08	ug/l	91
30) Chloroform	5.102	83	3067	0.83	ug/l	91
40) Benzene	5.787	78	9135	1.22	ug/l	93
44) Trichloroethene	6.549	130	122576	61.29	ug/l	95
48) Methyl methacrylate	6.967	41	1731	0.57	ug/l #	85
52) Toluene	7.976	92	189497	40.91	ug/l	97
95) Naphthalene	14.095	128	6617	3.06	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
Data File : VU042397.D
Acq On : 22 Feb 2021 16:06
Operator : SY/MD
Sample : M1457-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GWBR1-70-85-021821

Quant Time: Feb 22 16:28:33 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

