

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
 Data File : VU042401.D
 Acq On : 22 Feb 2021 17:34
 Operator : SY/MD
 Sample : PB134750TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB134750TB

Quant Time: Feb 23 01:04:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	178697	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	267262	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	241699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	110633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	128977	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
35) Dibromofluoromethane	5.298	113	90705	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
50) Toluene-d8	7.906	98	332242	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	10.639	95	117776	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds					Qvalue	
16) Acetone	2.646	43	12526	7.45	ug/l	96
20) Methylene Chloride	3.057	84	4079	2.10	ug/l	91
43) Isopropyl Acetate	5.919	43	20385	3.38	ug/l	96
95) Naphthalene	14.095	128	5172	2.88	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
Data File : VU042401.D
Acq On : 22 Feb 2021 17:34
Operator : SY/MD
Sample : PB134750TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB134750TB

Quant Time: Feb 23 01:04:45 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

