

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
 Data File : VU042403.D
 Acq On : 22 Feb 2021 18:20
 Operator : SY/MD
 Sample : PB134750ZHE#02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB134750ZHE#02

Quant Time: Feb 23 01:05:19 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	178739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	270655	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	243970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	119527	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	130347	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
35) Dibromofluoromethane	5.298	113	91248	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
50) Toluene-d8	7.906	98	332689	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	10.639	95	121614	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
						Qvalue
16) Acetone	2.646	43	12425	7.39	ug/l	92
18) Methyl Acetate	2.964	43	5197	1.70	ug/l	98
20) Methylene Chloride	3.054	84	3877	2.00	ug/l	96
43) Isopropyl Acetate	5.922	43	19646	3.21	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
Data File : VU042403.D
Acq On : 22 Feb 2021 18:20
Operator : SY/MD
Sample : PB134750ZHE#02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB134750ZHE#02

Quant Time: Feb 23 01:05:19 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

