

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
 Data File : VU042404.D
 Acq On : 22 Feb 2021 18:43
 Operator : SY/MD
 Sample : PB134750ZHE#03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB134750ZHE#03

Quant Time: Feb 23 01:05:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	173427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	264625	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	240422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	116168	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	127071	48.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.62%	
35) Dibromofluoromethane	5.298	113	90227	48.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.62%	
50) Toluene-d8	7.906	98	325724	48.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.38%	
62) 4-Bromofluorobenzene	10.639	95	120596	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	
Target Compounds						
						Qvalue
16) Acetone	2.645	43	13775	8.44	ug/l	98
18) Methyl Acetate	2.964	43	5094	1.72	ug/l #	90
20) Methylene Chloride	3.054	84	3793	2.02	ug/l	88
43) Isopropyl Acetate	5.922	43	20146	3.37	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
Data File : VU042404.D
Acq On : 22 Feb 2021 18:43
Operator : SY/MD
Sample : PB134750ZHE#03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB134750ZHE#03

Quant Time: Feb 23 01:05:37 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

