

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
 Data File : VU042408.D
 Acq On : 22 Feb 2021 20:15
 Operator : SY/MD
 Sample : PB134750ZHE#07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB134750ZHE#07

Quant Time: Feb 23 01:06:42 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 19 02:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	179007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	271574	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	248818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	121489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	131032	48.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.54%	
35) Dibromofluoromethane	5.298	113	90807	47.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.74%	
50) Toluene-d8	7.906	98	333840	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
62) 4-Bromofluorobenzene	10.639	95	123207	45.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.06%	
Target Compounds						
						Qvalue
16) Acetone	2.646	43	12202	7.25	ug/l	99
18) Methyl Acetate	2.961	43	6924	2.26	ug/l	95
20) Methylene Chloride	3.057	84	3615	1.86	ug/l	89
43) Isopropyl Acetate	5.919	43	19612	3.20	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022221\
Data File : VU042408.D
Acq On : 22 Feb 2021 20:15
Operator : SY/MD
Sample : PB134750ZHE#07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB134750ZHE#07

Quant Time: Feb 23 01:06:42 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U021721W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 19 02:55:52 2021
Response via : Initial Calibration

