

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022222\
 Data File : VU047232.D
 Acq On : 22 Feb 2022 15:21
 Operator : SY/MD
 Sample : VU0222WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK069

Quant Time: Feb 23 03:39:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 18 00:54:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	374193	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	363893	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	208418	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	161670	47.572	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.140%	
7) Chloroethane-d5	1.919	69	131293	52.116	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	104.240%	
11) 1,1-Dichloroethene-d2	2.571	63	214839	38.148	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.300%	
21) 2-Butanone-d5	4.636	46	281205	106.692	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	106.690%	
24) Chloroform-d	5.070	84	266799	48.359	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.720%	
26) 1,2-Dichloroethane-d4	5.706	65	172609	48.744	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.480%	
32) Benzene-d6	5.732	84	569900	47.108	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.220%	
36) 1,2-Dichloropropane-d6	6.693	67	179379	45.722	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.440%	
41) Toluene-d8	7.899	98	541579	47.017	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.040%	
43) trans-1,3-Dichloroprop...	8.182	79	90133	43.839	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.680%	
47) 2-Hexanone-d5	8.636	63	228396	101.617	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	101.620%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	262928	45.961	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	91.920%	
66) 1,2-Dichlorobenzene-d4	12.195	152	214719	43.439	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.880%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022222\
Data File : VU047232.D
Acq On : 22 Feb 2022 15:21
Operator : SY/MD
Sample : VU02222WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK069

Quant Time: Feb 23 03:39:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 18 00:54:38 2022
Response via : Initial Calibration

