

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022322\
 Data File : VU047246.D
 Acq On : 23 Feb 2022 12:46
 Operator : SY/MD
 Sample : N1637-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L10-M-022222

Quant Time: Feb 24 05:58:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.118	96	51182	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	16579	0.836	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	16856	0.816	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.000%	
Target Compounds						
13) Acetone	2.652	43	5434	0.494	ug/l	100
15) Methylene Chloride	3.051	84	3379	0.249	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022322\
Data File : VU047246.D
Acq On : 23 Feb 2022 12:46
Operator : SY/MD
Sample : N1637-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L10-M-022222

Quant Time: Feb 24 05:58:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 15 03:17:28 2022
Response via : Initial Calibration

