

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022322\
 Data File : VU047241.D
 Acq On : 23 Feb 2022 10:01
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 24 05:56:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.333	0.325	2.4	85	0.00
3 t	Chloromethane	0.342	0.391	-14.3	100	0.00
4 Rt	Vinyl Chloride	0.338	0.318	5.9	83	0.00
5 T	Bromomethane	0.221	0.158	28.5	68	0.00
6 T	Chloroethane	0.194	0.194	0.0	89	0.00
7 T	Trichlorofluoromethane	0.398	0.420	-5.5	92	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.230	0.237	-3.0	90	0.00
9 Rt	1,1-Dichloroethene	0.236	0.234	0.8	90	0.00
10 t	Iodomethane	0.223	0.200	10.3	68	0.00
11 t	Allyl Chloride	0.368	0.336	8.7	80	0.00
12 t	Acrylonitrile	0.069	0.066	4.3	99	0.00
13 T	Acetone	0.040	0.052	-30.0	128	0.02
14 T	Carbon Disulfide	0.800	0.763	4.6	84	0.00
15 RT	Methylene Chloride	0.265	0.262	1.1	89	0.00
16 RT	trans-1,2-Dichloroethene	0.249	0.253	-1.6	89	0.00
17 t	1,1-Dichloroethane	0.457	0.438	4.2	84	0.00
18 T	2-Butanone	0.070	0.081	-15.7	107	0.01
19	Cyclohexane	0.423	0.393	7.1	81	0.00
20	Methylcyclohexane	0.467	0.445	4.7	83	0.00
21 T	2,2-Dichloropropane	0.408	0.403	1.2	86	0.00
22 RT	cis-1,2-Dichloroethene	0.287	0.280	2.4	87	0.00
23 t	Diethyl Ether	0.191	0.187	2.1	87	0.00
24 t	tert-Butyl Alcohol	0.012	0.029	-141.7#	211#	0.08
25 t	Methyl tert-Butyl Ether	0.663	0.644	2.9	87	0.00
26 t	Bromochloromethane	0.129	0.131	-1.6	89	0.00
27 t	Chloroform	0.464	0.459	1.1	90	0.00
28 RT	1,1,1-Trichloroethane	0.402	0.398	1.0	87	0.00
29 T	1,1-Dichloropropene	0.368	0.356	3.3	85	0.00
30 RT	Carbon Tetrachloride	0.353	0.351	0.6	86	0.00
31 t	Isopropyl Ether	0.731	0.699	4.4	83	0.00
32	Ethyl-t-butyl ether	0.743	0.700	5.8	82	0.00
33	Tert-Amyl methyl ether	0.684	0.649	5.1	83	0.00
34 t	Propionitrile	0.019	0.026	-36.8#	129	0.02
35 RT	Benzene	1.030	1.020	1.0	86	0.00
36 RT	1,2-Dichloroethane	0.312	0.307	1.6	86	0.00
37 RT	Trichloroethene	0.285	0.284	0.4	87	0.00
38 Rt	1,2-Dichloropropane	0.273	0.262	4.0	84	0.00
39 t	Methacrylonitrile	0.094	0.094	0.0	92	0.00
40 t	Methyl acrylate	0.159	0.186	-17.0	108	0.00
41 t	Tetrahydrofuran	0.048	0.050	-4.2	108	0.00
42 t	1-Chlorobutane	0.494	0.474	4.0	84	0.00
43 T	Dibromomethane	0.147	0.148	-0.7	87	0.00
44 T	Bromodichloromethane	0.348	0.343	1.4	85	0.00
45 T	4-Methyl-2-Pentanone	0.199	0.187	6.0	81	0.00
46 t	t-1,4-Dichloro-2-butene	0.087	0.096	-10.3	117	0.00
47 t	Methyl methacrylate	0.154	0.149	3.2	87	0.00
48 t	Ethyl methacrylate	0.313	0.298	4.8	80	0.00
49 Rt	Toluene	0.664	0.654	1.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022322\
 Data File : VU047241.D
 Acq On : 23 Feb 2022 10:01
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Feb 24 05:56:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021422DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Feb 15 03:17:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.376	0.363	3.5	82	0.00
51 T	cis-1,3-Dichloropropene	0.422	0.409	3.1	83	0.00
52 RT	1,1,2-Trichloroethane	0.210	0.209	0.5	87	0.00
53 t	1,3-Dichloropropane	0.375	0.355	5.3	82	0.00
54 t	2-Hexanone	0.148	0.138	6.8	82	0.00
55 t	Dibromochloromethane	0.255	0.252	1.2	83	0.00
56 T	1,2-Dibromoethane	0.212	0.209	1.4	84	0.00
57 S	4-Bromofluorobenzene	0.387	0.373	3.6	88	0.00
58 RT	Tetrachloroethene	0.263	0.254	3.4	84	0.00
59 Rt	Chlorobenzene	0.741	0.719	3.0	84	0.00
60 T	1,1,1,2-Tetrachloroethane	0.255	0.259	-1.6	86	0.00
61 t	Pentachloroethane	0.193	0.210	-8.8	90	0.00
62 t	Hexachloroethane	0.209	0.209	0.0	83	0.00
63 Rt	Ethyl Benzene	1.302	1.261	3.1	84	0.00
64 RT	m/p-Xylenes	0.496	0.483	2.6	83	0.00
65 RT	o-Xylene	0.481	0.468	2.7	83	0.00
66 RT	Styrene	0.816	0.809	0.9	84	0.00
67 t	Bromoform	0.162	0.164	-1.2	85	0.00
68 S	1,2-Dichlorobenzene-d4	0.403	0.433	-7.4	98	0.00
69 T	Isopropylbenzene	1.293	1.250	3.3	83	0.00
70 T	1,1,2,2-Tetrachloroethane	0.284	0.285	-0.4	87	0.00
71 T	1,2,3-Trichloropropane	0.239	0.211	11.7	72	0.00
72 t	Bromobenzene	0.316	0.321	-1.6	87	0.00
73 t	n-propylbenzene	0.356	0.351	1.4	84	0.00
74 t	2-Chlorotoluene	0.304	0.303	0.3	85	0.00
75 t	1,3,5-Trimethylbenzene	1.061	1.066	-0.5	85	0.00
76 t	4-Chlorotoluene	0.319	0.321	-0.6	86	0.00
77 t	tert-Butylbenzene	1.064	1.066	-0.2	85	0.00
78 t	1,2,4-Trimethylbenzene	1.082	1.089	-0.6	87	0.00
79 t	sec-Butylbenzene	1.413	1.407	0.4	84	0.00
80	Nitrobenzene	0.014	0.009	35.7#	54	0.00
81 t	p-Isopropyltoluene	1.196	1.208	-1.0	86	0.00
82 t	1,3-Dichlorobenzene	0.615	0.625	-1.6	87	0.00
83 Rt	1,4-Dichlorobenzene	0.619	0.628	-1.5	88	0.00
84 t	n-Butylbenzene	1.148	1.191	-3.7	87	0.00
85 Rt	1,2-Dichlorobenzene	0.592	0.607	-2.5	89	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.052	0.054	-3.8	91	0.00
87 Rt	1,2,4-Trichlorobenzene	0.465	0.511	-9.9	94	0.00
88 t	Hexachlorobutadiene	0.238	0.258	-8.4	93	0.00
89 t	Naphthalene	0.952	0.996	-4.6	91	0.00
90 t	1,2,3-Trichlorobenzene	0.417	0.467	-12.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0