

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022519\
 Data File : VU029662.D
 Acq On : 25 Feb 2019 16:50
 Operator : JC/SP
 Sample : K1600-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF7Y5

Quant Time: Feb 26 01:19:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 26 00:12:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	455813	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	438035	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	213201	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	139641	41.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.62%
7) Chloroethane-d5	1.68	69	126065	42.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.92%
11) 1,1-Dichloroethene-d2	2.27	63	203489	31.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.34%
21) 2-Butanone-d5	4.17	46	208182	102.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.42%
24) Chloroform-d	4.65	84	247304	45.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.30%
26) 1,2-Dichloroethane-d4	5.31	65	162288	47.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.64%
32) Benzene-d6	5.34	84	562457	51.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.06%
36) 1,2-Dichloropropane-d6	6.33	67	171092	51.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.18%
41) Toluene-d8	7.56	98	528335	46.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.94%
43) trans-1,3-Dichloropropene-	7.85	79	78318	44.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.06%
47) 2-Hexanone-d5	8.31	63	163573	88.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.58%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	260435	46.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.96%
64) 1,2-Dichlorobenzene-d4	11.86	152	221800	48.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022519\
Data File : VU029662.D
Acq On : 25 Feb 2019 16:50
Operator : JC/SP
Sample : K1600-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF7Y5

Quant Time: Feb 26 01:19:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 26 00:12:24 2019
Response via : Initial Calibration

