

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022520\
 Data File : VU036898.D
 Acq On : 25 Feb 2020 11:02
 Operator : JC/MD
 Sample : VU0225WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK58

Quant Time: Feb 26 01:43:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 26 01:28:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	266702	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	249044	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	95855	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	104294	46.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.14%
7) Chloroethane-d5	1.93	69	89467	48.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.56%
11) 1,1-Dichloroethene-d2	2.59	63	135133	36.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.06%
21) 2-Butanone-d5	4.67	46	153128	101.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.93%
24) Chloroform-d	5.11	84	165953	48.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.04%
26) 1,2-Dichloroethane-d4	5.74	65	114710	49.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.12%
32) Benzene-d6	5.76	84	365459	50.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.48%
36) 1,2-Dichloropropane-d6	6.73	67	119938	51.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.80%
41) Toluene-d8	7.92	98	332980	49.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.68%
43) trans-1,3-Dichloropropene-	8.20	79	52063	48.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.02%
47) 2-Hexanone-d5	8.66	63	114109	100.02	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.02%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	161684	49.15	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.30%
64) 1,2-Dichlorobenzene-d4	12.21	152	99427	53.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022520\
Data File : VU036898.D
Acq On : 25 Feb 2020 11:02
Operator : JC/MD
Sample : VU0225WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK58

Quant Time: Feb 26 01:43:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 26 01:28:52 2020
Response via : Initial Calibration

