

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022520\
 Data File : VU036902.D
 Acq On : 25 Feb 2020 12:59
 Operator : JC/MD
 Sample : L1608-08RE
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR2RE

Quant Time: Feb 26 03:11:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 26 01:28:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	201871	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	188436	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	69150	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	104120	61.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	122.84%
7) Chloroethane-d5	1.93	69	90001	64.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	128.32%
11) 1,1-Dichloroethene-d2	2.59	63	133480	47.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.04%
21) 2-Butanone-d5	4.67	46	159669	140.41	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	140.41%#
24) Chloroform-d	5.11	84	165461	63.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	127.84%#
26) 1,2-Dichloroethane-d4	5.74	65	115544	65.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	131.90%#
32) Benzene-d6	5.76	84	366705	67.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	134.58%#
36) 1,2-Dichloropropane-d6	6.73	67	118284	67.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	134.00%#
41) Toluene-d8	7.92	98	329264	65.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	130.26%#
43) trans-1,3-Dichloropropene-	8.20	79	51199	62.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	124.80%
47) 2-Hexanone-d5	8.66	63	119184	138.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	138.07%#
57) 1,1,2,2-Tetrachloroethane-	10.77	84	163438	65.66	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	131.32%#
64) 1,2-Dichlorobenzene-d4	12.21	152	94301	69.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	139.52%#

Target Compounds

					Ovalue
13) Acetone	2.66	43	3220	2.753 ug/L	93
20) cis-1,2-Dichloroethene	4.71	96	7236	4.620 ug/L	94
30) 1,1,1-Trichloroethane	5.36	97	52444	25.958 ug/L	97
34) Trichloroethene	6.57	95	381509	260.746 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022520\
Data File : VU036902.D
Acq On : 25 Feb 2020 12:59
Operator : JC/MD
Sample : L1608-08RE
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR2RE

Quant Time: Feb 26 03:11:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 26 01:28:52 2020
Response via : Initial Calibration

