

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
 Data File : VU042426.D
 Acq On : 25 Feb 2021 12:36
 Operator : SY/MD
 Sample : VSTD0.135
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1205

Quant Time: Feb 26 06:09:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 26 06:08:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.226	114	4234	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	3843	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	2151	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	303	0.07	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	452	0.08	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.658	67	317	0.07	ug/L	0.00
8) Toluene-d8	7.846	98	963	0.11	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.697	84	271	0.06	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.606	62	349	0.06	ug/L #	79
6) Trichloroethene	6.508	95	341	0.07	ug/L	97
9) 1,2-Dibromoethane	8.866	107	232	0.07	ug/L #	89
12) 1,2,3-Trichloropropane	10.748	75	238	0.07	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.924	75	45	0.07	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
Data File : VU042426.D
Acq On : 25 Feb 2021 12:36
Operator : SY/MD
Sample : VSTD0.135
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1205

Quant Time: Feb 26 06:09:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 26 06:08:59 2021
Response via : Initial Calibration

