

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
 Data File : VU042429.D
 Acq On : 25 Feb 2021 13:48
 Operator : SY/MD
 Sample : VSTD00232
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002202

Quant Time: Feb 26 06:10:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 26 06:08:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	4239	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	3798	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	2136	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	6701	1.56	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.679	65	7291	1.30	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	6958	1.58	ug/L	-0.02
8) Toluene-d8	7.845	98	18187	2.03	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.696	84	5735	1.32	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	6835	1.26	ug/L	96
6) Trichloroethene	6.508	95	6714	1.49	ug/L	99
9) 1,2-Dibromoethane	8.866	107	4712	1.48	ug/L #	98
12) 1,2,3-Trichloropropane	10.748	75	4569	1.44	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.924	75	918	1.41	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
Data File : VU042429.D
Acq On : 25 Feb 2021 13:48
Operator : SY/MD
Sample : VSTD00232
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002202

Quant Time: Feb 26 06:10:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 26 06:08:59 2021
Response via : Initial Calibration

