

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
 Data File : VU042434.D
 Acq On : 25 Feb 2021 17:14
 Operator : SY/MD
 Sample : M1496-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 X3164

Quant Time: Feb 26 06:25:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 26 06:20:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	3878	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	3498	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.758	152	1955	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	1571	0.52	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	104.00%		
4) 1,2-Dichloroethane-d4	5.679	65	1824	0.46	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	92.00%		
7) 1,2-Dichloropropane-d6	6.658	67	1635	0.53	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	106.00%		
8) Toluene-d8	7.845	98	4302	0.49	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.00%		
10) 1,1,2,2-Tetrachloroeth...	10.696	84	1340	0.52	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	104.00%		
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	657	0.21	ug/L	89
6) Trichloroethene	6.507	95	3052	0.97	ug/L	99
9) 1,2-Dibromoethane	8.866	107	3643	1.71	ug/L #	98
12) 1,2,3-Trichloropropane	10.747	75	1150	0.55	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.924	75	755	1.86	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022521\
Data File : VU042434.D
Acq On : 25 Feb 2021 17:14
Operator : SY/MD
Sample : M1496-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
X3164

Quant Time: Feb 26 06:25:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 26 06:20:15 2021
Response via : Initial Calibration

