

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022620\
 Data File : VU036925.D
 Acq On : 26 Feb 2020 19:06
 Operator : JC/MD
 Sample : L1687-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AT3

Quant Time: Feb 27 06:30:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 27 04:47:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	245092	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	230854	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	87848	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	89563	43.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.02%
7) Chloroethane-d5	1.93	69	81940	48.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.22%
11) 1,1-Dichloroethene-d2	2.59	63	121901	35.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.74%
21) 2-Butanone-d5	4.67	46	148418	107.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.50%
24) Chloroform-d	5.11	84	154606	49.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.38%
26) 1,2-Dichloroethane-d4	5.74	65	107807	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.36%
32) Benzene-d6	5.76	84	336388	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.78%
36) 1,2-Dichloropropane-d6	6.72	67	109588	50.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.34%
41) Toluene-d8	7.92	98	298237	48.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.30%
43) trans-1,3-Dichloropropene-	8.20	79	46600	46.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.72%
47) 2-Hexanone-d5	8.66	63	113089	106.94	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.94%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	153242	50.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.50%
64) 1,2-Dichlorobenzene-d4	12.21	152	90620	52.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.54%

Target Compounds

					Ovalue
10) 1,1,2-Trichloro-1,2,2-trif	2.60	101	4314	2.756 ug/L #	78
12) 1,1-Dichloroethene	2.61	96	10152	6.413 ug/L #	1
13) Acetone	2.66	43	2552	1.797 ug/L	90
19) 1,1-Dichloroethane	3.91	63	3161	0.957 ug/L	93
20) cis-1,2-Dichloroethene	4.71	96	28427	14.950 ug/L	98
30) 1,1,1-Trichloroethane	5.36	97	205523	83.036 ug/L	100
34) Trichloroethene	6.57	95	913050	509.370 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022620\
Data File : VU036925.D
Acq On : 26 Feb 2020 19:06
Operator : JC/MD
Sample : L1687-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AT3

Quant Time: Feb 27 06:30:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 27 04:47:27 2020
Response via : Initial Calibration

