

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022620\
 Data File : VU036928.D
 Acq On : 26 Feb 2020 20:14
 Operator : JC/MD
 Sample : L1687-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AT6

Quant Time: Feb 27 06:53:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 27 04:47:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	256634	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	243229	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	93088	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	88419	41.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.06%
7) Chloroethane-d5	1.93	69	80185	44.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.94%
11) 1,1-Dichloroethene-d2	2.59	63	114855	31.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.64%
21) 2-Butanone-d5	4.67	46	151407	104.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.73%
24) Chloroform-d	5.11	84	151974	46.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.36%
26) 1,2-Dichloroethane-d4	5.74	65	105113	47.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.40%
32) Benzene-d6	5.76	84	332340	47.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.50%
36) 1,2-Dichloropropane-d6	6.73	67	110051	48.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.58%
41) Toluene-d8	7.93	98	297318	45.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.12%
43) trans-1,3-Dichloropropene-	8.20	79	45835	43.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.56%
47) 2-Hexanone-d5	8.66	63	113108	101.51	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.51%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	154537	48.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	91721	50.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds

					Ovalue
13) Acetone	2.66	43	3243	2.181 ug/L	97
20) cis-1,2-Dichloroethene	4.71	96	6155	3.091 ug/L	98
30) 1,1,1-Trichloroethane	5.36	97	8641	3.314 ug/L	99
34) Trichloroethene	6.57	95	69527	36.814 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022620\
Data File : VU036928.D
Acq On : 26 Feb 2020 20:14
Operator : JC/MD
Sample : L1687-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AT6

Quant Time: Feb 27 06:53:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 27 04:47:27 2020
Response via : Initial Calibration

