

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022620\
 Data File : VU036909.D
 Acq On : 26 Feb 2020 11:09
 Operator : JC/MD
 Sample : VU0226WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK59

Quant Time: Feb 27 04:55:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 27 04:47:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	259574	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	244194	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	90444	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	98997	45.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.84%
7) Chloroethane-d5	1.93	69	88774	49.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.44%
11) 1,1-Dichloroethene-d2	2.59	63	127038	34.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.60%
21) 2-Butanone-d5	4.67	46	157941	108.02	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.02%
24) Chloroform-d	5.11	84	162903	48.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.88%
26) 1,2-Dichloroethane-d4	5.74	65	114461	50.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.62%
32) Benzene-d6	5.76	84	360575	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.12%
36) 1,2-Dichloropropane-d6	6.73	67	118752	51.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.82%
41) Toluene-d8	7.92	98	322634	49.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.50%
43) trans-1,3-Dichloropropene-	8.20	79	51580	48.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.02%
47) 2-Hexanone-d5	8.66	63	118162	105.63	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.63%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	163072	50.56	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.12%
64) 1,2-Dichlorobenzene-d4	12.21	152	94226	53.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022620\
Data File : VU036909.D
Acq On : 26 Feb 2020 11:09
Operator : JC/MD
Sample : VU0226WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK59

Quant Time: Feb 27 04:55:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 27 04:47:27 2020
Response via : Initial Calibration

