

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022621\
 Data File : VU042438.D
 Acq On : 26 Feb 2021 09:21
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5211

Quant Time: Feb 26 21:48:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 26 06:20:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	4353	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	3721	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	2003	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	1611	0.47	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	94.00%		
4) 1,2-Dichloroethane-d4	5.680	65	2003	0.45	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	90.00%		
7) 1,2-Dichloropropane-d6	6.658	67	1672	0.51	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	102.00%		
8) Toluene-d8	7.845	98	4512	0.48	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	96.00%		
10) 1,1,2,2-Tetrachloroeth...	10.696	84	1432	0.52	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	104.00%		
Target Compounds						
3) Vinyl chloride	1.606	62	1617	0.45	ug/L	100
6) Trichloroethene	6.508	95	1596	0.48	ug/L	99
9) 1,2-Dibromoethane	8.866	107	1123	0.50	ug/L	99
12) 1,2,3-Trichloropropane	10.748	75	1086	0.50	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.924	75	223	0.53	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022621\
Data File : VU042438.D
Acq On : 26 Feb 2021 09:21
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5211

Quant Time: Feb 26 21:48:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 26 06:20:15 2021
Response via : Initial Calibration

