

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022621\
 Data File : VU042439.D
 Acq On : 26 Feb 2021 09:55
 Operator : SY/MD
 Sample : VU0226WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK205

Quant Time: Feb 26 21:52:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 26 21:52:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	4083	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	3656	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	1956	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.606	65	1703	0.53	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.00%			
4) 1,2-Dichloroethane-d4	5.680	65	1949	0.47	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 94.00%			
7) 1,2-Dichloropropane-d6	6.658	67	1754	0.54	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.00%			
8) Toluene-d8	7.845	98	4583	0.50	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 100.00%			
10) 1,1,2,2-Tetrachloroeth...	10.696	84	1465	0.54	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 108.00%			

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU022621\
Data File : VU042439.D
Acq On : 26 Feb 2021 09:55
Operator : SY/MD
Sample : VU0226WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK205

Quant Time: Feb 26 21:52:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM022521.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 26 21:52:13 2021
Response via : Initial Calibration

